

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012821\
 Data File : P0075284.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2021 20:19
 Operator : AJ/MA
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO012821AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 07:53:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 07:43:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.923	3.995	5097125	2429097	50.366	50.541
2)	SA Decachlor...	10.936	9.244	5114591	2567921	53.045	52.927
Target Compounds							
21)	L5 AR-1248-1	6.246	5.239	1154766	572261	496.113	476.747
22)	L5 AR-1248-2	6.542	5.501	1497966	792314	488.219	480.358
23)	L5 AR-1248-3	6.760	5.545	1764125	801851	490.665	481.887
24)	L5 AR-1248-4	7.190	5.727	2158601	950982	487.194	483.182
25)	L5 AR-1248-5	7.229	6.146	2055831	1001586	487.015	479.505



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012821\
Data File : PO075284.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jan 2021 20:19
Operator : AJ/MA
Sample : AR1248ICV500
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO012821AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 07:53:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 07:43:22 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 5097125
Conc: 50.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO012821AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.995 min
Delta R.T.: -0.001 min
Response: 2429097
Conc: 50.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.936 min
Delta R.T.: -0.002 min
Response: 5114591
Conc: 53.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.244 min
Delta R.T.: -0.002 min
Response: 2567921
Conc: 52.93 ng/ml



#21 AR-1248-1


R.T.: 6.246 min
Delta R.T.: -0.001 min
Response: 1154766
Conc: 496.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO012821AR1248




#21 AR-1248-1

R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 572261
Conc: 476.75 ng/ml




#22 AR-1248-2

R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 1497966
Conc: 488.22 ng/ml



#22 AR-1248-2


R.T.: 5.501 min
Delta R.T.: 0.000 min
Response: 792314
Conc: 480.36 ng/ml



#23 AR-1248-3


R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 1764125
Conc: 490.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO012821AR1248




#23 AR-1248-3

R.T.: 5.545 min
Delta R.T.: -0.001 min
Response: 801851
Conc: 481.89 ng/ml




#24 AR-1248-4

R.T.: 7.190 min
Delta R.T.: 0.000 min
Response: 2158601
Conc: 487.19 ng/ml



#24 AR-1248-4


R.T.: 5.727 min
Delta R.T.: -0.001 min
Response: 950982
Conc: 483.18 ng/ml



#25 AR-1248-5

R.T.: 7.229 min
Delta R.T.: 0.000 min
Response: 2055831
Conc: 487.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO012821AR1248



#25 AR-1248-5

R.T.: 6.146 min
Delta R.T.: -0.001 min
Response: 1001586
Conc: 479.50 ng/ml