

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012821\
 Data File : P0075299.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Jan 2021 1:27
 Operator : AJ/MA
 Sample : M1246-20
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 250EM-013-0204

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 08:01:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 07:43:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.922	3.995	2178144	1044004	21.523	21.722
2) SA Decachlor...	10.934	9.245	2031856	1057285	21.073	21.792
Target Compounds						
36) L8 AR-1262-1	8.571	7.342	206044	57799	23.611	26.498
37) L8 AR-1262-2	9.127	7.889	199946	112164	13.222	14.626
38) L8 AR-1262-3	9.437	8.173	387912	224926	37.132	67.967 #
39) L8 AR-1262-4	9.533	8.238	367260	209724	49.689	36.526 #
40) L8 AR-1262-5	10.195	8.731	198425	123293	38.262	44.820



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012821\
Data File : P0075299.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Jan 2021 1:27
Operator : AJ/MA
Sample : M1246-20
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
250EM-013-0204

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 08:01:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 07:43:22 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.922 min
Delta R.T.: -0.002 min
Response: 2178144
Conc: 21.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
250EM-013-0204




#1 Tetrachloro-m-xylene

R.T.: 3.995 min
Delta R.T.: -0.002 min
Response: 1044004
Conc: 21.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.934 min
Delta R.T.: -0.003 min
Response: 2031856
Conc: 21.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.245 min
Delta R.T.: -0.001 min
Response: 1057285
Conc: 21.79 ng/ml



#36 AR-1262-1


R.T.: 8.571 min
Delta R.T.: 0.000 min
Response: 206044
Conc: 23.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
250EM-013-0204




#36 AR-1262-1

R.T.: 7.342 min
Delta R.T.: 0.001 min
Response: 57799
Conc: 26.50 ng/ml




#37 AR-1262-2

R.T.: 9.127 min
Delta R.T.: -0.002 min
Response: 199946
Conc: 13.22 ng/ml



#37 AR-1262-2


R.T.: 7.889 min
Delta R.T.: 0.000 min
Response: 112164
Conc: 14.63 ng/ml



#38 AR-1262-3


R.T.: 9.437 min
Delta R.T.: -0.002 min
Response: 387912
Conc: 37.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
250EM-013-0204




#38 AR-1262-3

R.T.: 8.173 min
Delta R.T.: 0.000 min
Response: 224926
Conc: 67.97 ng/ml




#39 AR-1262-4

R.T.: 9.533 min
Delta R.T.: 0.002 min
Response: 367260
Conc: 49.69 ng/ml



#39 AR-1262-4


R.T.: 8.238 min
Delta R.T.: 0.000 min
Response: 209724
Conc: 36.53 ng/ml



#40 AR-1262-5

R.T.: 10.195 min
Delta R.T.: -0.001 min
Response: 198425
Conc: 38.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
250EM-013-0204



#40 AR-1262-5

R.T.: 8.731 min
Delta R.T.: -0.001 min
Response: 123293
Conc: 44.82 ng/ml