

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012822\
 Data File : P0084460.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2022 16:25
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 22:56:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 28 07:54:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	6.518	5.303	2953475	998640	51.237	53.616
2) SA Decachlor...	13.747	11.301	3668580	1328254	43.740	46.053
Target Compounds						
3) L1 AR-1016-1	7.969	6.770	1120451	433459	486.359	520.520
4) L1 AR-1016-2	7.994	6.792	1557337	623965	453.236	522.816
5) L1 AR-1016-3	8.068	7.010	1033779	328164	463.669	511.927
6) L1 AR-1016-4	8.183	7.063	831117	270619	490.760	509.096
7) L1 AR-1016-5	8.517	7.318	940661	364886	500.132	533.918
31) L7 AR-1260-1	9.754	8.494	1591352	658044	461.186	494.894
32) L7 AR-1260-2	10.043	8.696	1927153	798588	467.135	493.164
33) L7 AR-1260-3	10.480	8.865	1567778	749854	462.749	492.437
34) L7 AR-1260-4	10.775	9.368	1813798	648920	460.769	480.871
35) L7 AR-1260-5	11.196	9.618	3430844	1559570	459.809	479.558



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012822\
Data File : P0084460.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jan 2022 16:25
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 22:56:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012622.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 28 07:54:18 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 6.518 min
Delta R.T.: 0.000 min
Response: 2953475
Conc: 51.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 5.303 min
Delta R.T.: 0.000 min
Response: 998640
Conc: 53.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 13.747 min
Delta R.T.: 0.001 min
Response: 3668580
Conc: 43.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 11.301 min
Delta R.T.: 0.001 min
Response: 1328254
Conc: 46.05 ng/ml



#3 AR-1016-1


R.T.: 7.969 min
Delta R.T.: 0.000 min
Response: 1120451
Conc: 486.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 6.770 min
Delta R.T.: 0.000 min
Response: 433459
Conc: 520.52 ng/ml




#4 AR-1016-2

R.T.: 7.994 min
Delta R.T.: 0.000 min
Response: 1557337
Conc: 453.24 ng/ml



#4 AR-1016-2


R.T.: 6.792 min
Delta R.T.: 0.000 min
Response: 623965
Conc: 522.82 ng/ml



#5 AR-1016-3


R.T.: 8.068 min
Delta R.T.: 0.000 min
Response: 1033779
Conc: 463.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 7.010 min
Delta R.T.: 0.000 min
Response: 328164
Conc: 511.93 ng/ml




#6 AR-1016-4

R.T.: 8.183 min
Delta R.T.: 0.000 min
Response: 831117
Conc: 490.76 ng/ml



#6 AR-1016-4


R.T.: 7.063 min
Delta R.T.: 0.000 min
Response: 270619
Conc: 509.10 ng/ml



#7 AR-1016-5


R.T.: 8.517 min
Delta R.T.: 0.000 min
Response: 940661
Conc: 500.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 7.318 min
Delta R.T.: 0.000 min
Response: 364886
Conc: 533.92 ng/ml




#31 AR-1260-1

R.T.: 9.754 min
Delta R.T.: 0.000 min
Response: 1591352
Conc: 461.19 ng/ml



#31 AR-1260-1


R.T.: 8.494 min
Delta R.T.: 0.000 min
Response: 658044
Conc: 494.89 ng/ml



#32 AR-1260-2


R.T.: 10.043 min
Delta R.T.: 0.000 min
Response: 1927153
Conc: 467.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 798588
Conc: 493.16 ng/ml




#33 AR-1260-3

R.T.: 10.480 min
Delta R.T.: 0.000 min
Response: 1567778
Conc: 462.75 ng/ml



#33 AR-1260-3


R.T.: 8.865 min
Delta R.T.: 0.000 min
Response: 749854
Conc: 492.44 ng/ml



#34 AR-1260-4


R.T.: 10.775 min
Delta R.T.: -0.001 min
Response: 1813798
Conc: 460.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 9.368 min
Delta R.T.: 0.000 min
Response: 648920
Conc: 480.87 ng/ml



#35 AR-1260-5

R.T.: 11.196 min
Delta R.T.: 0.000 min
Response: 3430844
Conc: 459.81 ng/ml



#35 AR-1260-5

R.T.: 9.618 min
Delta R.T.: 0.000 min
Response: 1559570
Conc: 479.56 ng/ml