

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012921\
 Data File : P0075334.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Jan 2021 14:26
 Operator : AJ/MA
 Sample : M1246-21
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 03:16:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 09:09:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.923	3.995	2277910	1114007	22.509	23.174
2)	SA Decachlor...	10.938	9.244	2305153	1041299	23.892	21.452
Target Compounds							
3)	L1 AR-1016-1	6.246	5.258f	51987	59249	13.158	30.525 #
4)	L1 AR-1016-2	6.270	5.258	83092	59249	15.151	22.093 #
5)	L1 AR-1016-3	6.336	5.446	36344	20199	11.112	14.070 #
6)	L1 AR-1016-4	6.455	5.499	103255	50670	37.499	42.261
7)	L1 AR-1016-5	6.760	5.724	108225	26379	40.605	18.101 #
8)	L2 AR-1221-1	5.175	0.000	16398	0	13.715	N.D. #
9)	L2 AR-1221-2	5.277	4.342	32803	22658	37.157	48.712 #
10)	L2 AR-1221-3	5.373	0.000	38719	0	14.023	N.D. #
11)	L3 AR-1232-1	5.373	0.000	38719	0	17.232	N.D. #
12)	L3 AR-1232-2	5.948	5.258	79110	59249	69.699	49.582 #
13)	L3 AR-1232-3	6.270	5.446	83092	20199	34.695	31.819
14)	L3 AR-1232-4	6.455	5.543	103255	32841	87.605	59.200 #
15)	L3 AR-1232-5	6.543	5.724	81334	26379	101.251	44.289 #
16)	L4 AR-1242-1	6.246	5.258f	51987	59249	16.545	37.849 #
17)	L4 AR-1242-2	6.270	5.258	83092	59249	18.942	27.166 #
18)	L4 AR-1242-3	6.336	5.446	36344	20199	13.681	17.091
19)	L4 AR-1242-4	6.455	5.543	103255	32841	46.797	28.678 #
20)	L4 AR-1242-5	7.231	6.102	301730	193614	121.506	128.632
21)	L5 AR-1248-1	6.246	5.258f	51987	59249	22.335	49.360 #
22)	L5 AR-1248-2	6.543	5.499	81334	50670	26.509	30.720
23)	L5 AR-1248-3	6.760	5.543	108225	32841	30.101	19.736 #
24)	L5 AR-1248-4	7.190	5.724	252826	26379	57.062	13.403 #
25)	L5 AR-1248-5	7.231	6.145	301730	46889	71.478	22.448 #
26)	L6 AR-1254-1	7.161	6.102	285513	193614	66.246	62.513
27)	L6 AR-1254-2	7.390	6.262	519155	139355	77.001	51.895 #
28)	L6 AR-1254-3	7.773	6.678	702302	361008	101.840	84.664
29)	L6 AR-1254-4	8.070	6.916	463301	138882	76.635	47.649 #
30)	L6 AR-1254-5	8.499	7.341	1265804	285523	226.446	73.567 #
31)	L7 AR-1260-1	7.940	6.814	713738	199483	154.707	73.482 #
32)	L7 AR-1260-2	8.206	7.010	834474	209016	129.286	63.007 #
33)	L7 AR-1260-3	8.571	7.162	1367030	261066	250.520	87.841 #
34)	L7 AR-1260-4	8.817	7.640	1552299	463507	301.974	178.174 #
35)	L7 AR-1260-5	9.130	7.888	1398950	635454	118.543	100.414
36)	L8 AR-1262-1	8.571	7.341	1367030	285523	156.653	130.901
37)	L8 AR-1262-2	9.130	7.888	1398950	635454	92.511	82.860
38)	L8 AR-1262-3	9.438	8.173	2340188	940978	224.009	284.342 #
39)	L8 AR-1262-4	9.534	8.237	2338863	1152849	316.442	200.781 #
40)	L8 AR-1262-5	10.195	8.730	1290743	539345	248.894	196.065
41)	L9 AR-1268-1	9.438	8.173	2340188	940978	124.235	100.856

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012921\
 Data File : P0075334.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Jan 2021 14:26
 Operator : AJ/MA
 Sample : M1246-21
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 03:16:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 09:09:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.534	8.237	2338863	1152849	147.494	137.614
43) L9	AR-1268-3	9.758	8.442	490941	229929	35.327	31.760
44) L9	AR-1268-4	10.195	8.730	1290743	539345	219.109	171.190
45) L9	AR-1268-5	10.606	9.006	2469251	1105939	59.126	52.128



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012921\
Data File : PO075334.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Jan 2021 14:26
Operator : AJ/MA
Sample : M1246-21
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 03:16:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 09:09:46 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.923 min
Delta R.T.: -0.001 min
Response: 2277910
Conc: 22.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.995 min
Delta R.T.: -0.002 min
Response: 1114007
Conc: 23.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.938 min
Delta R.T.: 0.000 min
Response: 2305153
Conc: 23.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.244 min
Delta R.T.: -0.002 min
Response: 1041299
Conc: 21.45 ng/ml



#3 AR-1016-1


R.T.: 6.246 min
Delta R.T.: 0.000 min
Response: 51987
Conc: 13.16 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 5.258 min
Delta R.T.: 0.017 min
Response: 59249
Conc: 30.53 ng/ml




#4 AR-1016-2

R.T.: 6.270 min
Delta R.T.: 0.000 min
Response: 83092
Conc: 15.15 ng/ml



#4 AR-1016-2


R.T.: 5.258 min
Delta R.T.: -0.002 min
Response: 59249
Conc: 22.09 ng/ml



#5 AR-1016-3


R.T.: 6.336 min
Delta R.T.: -0.001 min
Response: 36344
Conc: 11.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 5.446 min
Delta R.T.: -0.003 min
Response: 20199
Conc: 14.07 ng/ml




#6 AR-1016-4

R.T.: 6.455 min
Delta R.T.: 0.011 min
Response: 103255
Conc: 37.50 ng/ml



#6 AR-1016-4


R.T.: 5.499 min
Delta R.T.: -0.003 min
Response: 50670
Conc: 42.26 ng/ml



#7 AR-1016-5


R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 108225
Conc: 40.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.724 min
Delta R.T.: -0.004 min
Response: 26379
Conc: 18.10 ng/ml




#8 AR-1221-1

R.T.: 5.175 min
Delta R.T.: 0.000 min
Response: 16398
Conc: 13.72 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 4.249 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 5.277 min
Delta R.T.: 0.002 min
Response: 32803
Conc: 37.16 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 4.342 min
Delta R.T.: -0.005 min
Response: 22658
Conc: 48.71 ng/ml




#10 AR-1221-3

R.T.: 5.373 min
Delta R.T.: 0.012 min
Response: 38719
Conc: 14.02 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 4.435 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 5.373 min
Delta R.T.: 0.013 min
Response: 38719
Conc: 17.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.435 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.948 min
Delta R.T.: -0.002 min
Response: 79110
Conc: 69.70 ng/ml



#12 AR-1232-2


R.T.: 5.258 min
Delta R.T.: -0.002 min
Response: 59249
Conc: 49.58 ng/ml



#13 AR-1232-3


R.T.: 6.270 min
Delta R.T.: 0.000 min
Response: 83092
Conc: 34.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 5.446 min
Delta R.T.: -0.002 min
Response: 20199
Conc: 31.82 ng/ml




#14 AR-1232-4

R.T.: 6.455 min
Delta R.T.: 0.011 min
Response: 103255
Conc: 87.60 ng/ml



#14 AR-1232-4


R.T.: 5.543 min
Delta R.T.: -0.002 min
Response: 32841
Conc: 59.20 ng/ml



#15 AR-1232-5


R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 81334
Conc: 101.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.724 min
Delta R.T.: -0.004 min
Response: 26379
Conc: 44.29 ng/ml




#16 AR-1242-1

R.T.: 6.246 min
Delta R.T.: 0.000 min
Response: 51987
Conc: 16.54 ng/ml



#16 AR-1242-1


R.T.: 5.258 min
Delta R.T.: 0.018 min
Response: 59249
Conc: 37.85 ng/ml



#17 AR-1242-2


R.T.: 6.270 min
Delta R.T.: -0.001 min
Response: 83092
Conc: 18.94 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 5.258 min
Delta R.T.: -0.001 min
Response: 59249
Conc: 27.17 ng/ml




#18 AR-1242-3

R.T.: 6.336 min
Delta R.T.: -0.003 min
Response: 36344
Conc: 13.68 ng/ml



#18 AR-1242-3


R.T.: 5.446 min
Delta R.T.: -0.002 min
Response: 20199
Conc: 17.09 ng/ml



#19 AR-1242-4


R.T.: 6.455 min
Delta R.T.: 0.010 min
Response: 103255
Conc: 46.80 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.543 min
Delta R.T.: -0.002 min
Response: 32841
Conc: 28.68 ng/ml




#20 AR-1242-5

R.T.: 7.231 min
Delta R.T.: 0.000 min
Response: 301730
Conc: 121.51 ng/ml



#20 AR-1242-5


R.T.: 6.102 min
Delta R.T.: -0.003 min
Response: 193614
Conc: 128.63 ng/ml



#21 AR-1248-1


R.T.: 6.246 min
Delta R.T.: 0.000 min
Response: 51987
Conc: 22.33 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 5.258 min
Delta R.T.: 0.018 min
Response: 59249
Conc: 49.36 ng/ml




#22 AR-1248-2

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 81334
Conc: 26.51 ng/ml



#22 AR-1248-2


R.T.: 5.499 min
Delta R.T.: -0.002 min
Response: 50670
Conc: 30.72 ng/ml



#23 AR-1248-3


R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 108225
Conc: 30.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.543 min
Delta R.T.: -0.003 min
Response: 32841
Conc: 19.74 ng/ml




#24 AR-1248-4

R.T.: 7.190 min
Delta R.T.: 0.000 min
Response: 252826
Conc: 57.06 ng/ml



#24 AR-1248-4


R.T.: 5.724 min
Delta R.T.: -0.004 min
Response: 26379
Conc: 13.40 ng/ml



#25 AR-1248-5


R.T.: 7.231 min
Delta R.T.: 0.000 min
Response: 301730
Conc: 71.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 6.145 min
Delta R.T.: -0.002 min
Response: 46889
Conc: 22.45 ng/ml




#26 AR-1254-1

R.T.: 7.161 min
Delta R.T.: 0.000 min
Response: 285513
Conc: 66.25 ng/ml



#26 AR-1254-1


R.T.: 6.102 min
Delta R.T.: -0.002 min
Response: 193614
Conc: 62.51 ng/ml



#27 AR-1254-2


R.T.: 7.390 min
Delta R.T.: 0.000 min
Response: 519155
Conc: 77.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 6.262 min
Delta R.T.: -0.002 min
Response: 139355
Conc: 51.89 ng/ml




#28 AR-1254-3

R.T.: 7.773 min
Delta R.T.: 0.000 min
Response: 702302
Conc: 101.84 ng/ml



#28 AR-1254-3


R.T.: 6.678 min
Delta R.T.: -0.001 min
Response: 361008
Conc: 84.66 ng/ml



#29 AR-1254-4


R.T.: 8.070 min
Delta R.T.: 0.000 min
Response: 463301
Conc: 76.64 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.916 min
Delta R.T.: -0.002 min
Response: 138882
Conc: 47.65 ng/ml




#30 AR-1254-5

R.T.: 8.499 min
Delta R.T.: 0.000 min
Response: 1265804
Conc: 226.45 ng/ml



#30 AR-1254-5


R.T.: 7.341 min
Delta R.T.: -0.003 min
Response: 285523
Conc: 73.57 ng/ml



#31 AR-1260-1


R.T.: 7.940 min
Delta R.T.: 0.000 min
Response: 713738
Conc: 154.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.814 min
Delta R.T.: -0.002 min
Response: 199483
Conc: 73.48 ng/ml




#32 AR-1260-2

R.T.: 8.206 min
Delta R.T.: 0.000 min
Response: 834474
Conc: 129.29 ng/ml



#32 AR-1260-2


R.T.: 7.010 min
Delta R.T.: -0.002 min
Response: 209016
Conc: 63.01 ng/ml



#33 AR-1260-3


R.T.: 8.571 min
Delta R.T.: 0.000 min
Response: 1367030
Conc: 250.52 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 7.162 min
Delta R.T.: -0.003 min
Response: 261066
Conc: 87.84 ng/ml




#34 AR-1260-4

R.T.: 8.817 min
Delta R.T.: 0.013 min
Response: 1552299
Conc: 301.97 ng/ml



#34 AR-1260-4


R.T.: 7.640 min
Delta R.T.: -0.003 min
Response: 463507
Conc: 178.17 ng/ml



#35 AR-1260-5


R.T.: 9.130 min
Delta R.T.: 0.000 min
Response: 1398950
Conc: 118.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.888 min
Delta R.T.: -0.004 min
Response: 635454
Conc: 100.41 ng/ml




#36 AR-1262-1

R.T.: 8.571 min
Delta R.T.: 0.000 min
Response: 1367030
Conc: 156.65 ng/ml



#36 AR-1262-1


R.T.: 7.341 min
Delta R.T.: 0.000 min
Response: 285523
Conc: 130.90 ng/ml



#37 AR-1262-2


R.T.: 9.130 min
Delta R.T.: 0.000 min
Response: 1398950
Conc: 92.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.888 min
Delta R.T.: -0.002 min
Response: 635454
Conc: 82.86 ng/ml




#38 AR-1262-3

R.T.: 9.438 min
Delta R.T.: 0.000 min
Response: 2340188
Conc: 224.01 ng/ml



#38 AR-1262-3


R.T.: 8.173 min
Delta R.T.: 0.000 min
Response: 940978
Conc: 284.34 ng/ml



#39 AR-1262-4


R.T.: 9.534 min
Delta R.T.: 0.003 min
Response: 2338863
Conc: 316.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 8.237 min
Delta R.T.: 0.000 min
Response: 1152849
Conc: 200.78 ng/ml




#40 AR-1262-5

R.T.: 10.195 min
Delta R.T.: 0.000 min
Response: 1290743
Conc: 248.89 ng/ml



#40 AR-1262-5


R.T.: 8.730 min
Delta R.T.: -0.001 min
Response: 539345
Conc: 196.06 ng/ml



#41 AR-1268-1


R.T.: 9.438 min
Delta R.T.: -0.001 min
Response: 2340188
Conc: 124.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 8.173 min
Delta R.T.: -0.002 min
Response: 940978
Conc: 100.86 ng/ml




#42 AR-1268-2

R.T.: 9.534 min
Delta R.T.: -0.002 min
Response: 2338863
Conc: 147.49 ng/ml



#42 AR-1268-2


R.T.: 8.237 min
Delta R.T.: -0.002 min
Response: 1152849
Conc: 137.61 ng/ml



#43 AR-1268-3


R.T.: 9.758 min
Delta R.T.: -0.003 min
Response: 490941
Conc: 35.33 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 8.442 min
Delta R.T.: -0.002 min
Response: 229929
Conc: 31.76 ng/ml




#44 AR-1268-4

R.T.: 10.195 min
Delta R.T.: -0.003 min
Response: 1290743
Conc: 219.11 ng/ml



#44 AR-1268-4


R.T.: 8.730 min
Delta R.T.: -0.001 min
Response: 539345
Conc: 171.19 ng/ml



#45 AR-1268-5

R.T.: 10.606 min
Delta R.T.: -0.002 min
Response: 2469251
Conc: 59.13 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 9.006 min
Delta R.T.: -0.003 min
Response: 1105939
Conc: 52.13 ng/ml