

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012921\
 Data File : P0075358.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Jan 2021 1:34
 Operator : AJ/MA
 Sample : M1246-05 10X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 250EM-007-0002

Manual Integrations
 APPROVED

Sohil
 2/1/2021 2:50:34 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 03:24:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 09:09:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.921	3.994	191171	101012	1.889	2.101
2) SA Decachlor...	10.932	9.242	7283147	3402662	75.487	70.098
Target Compounds						
36) L8 AR-1262-1	8.570	7.343	34836246	500782	3992.012	229.588 #
37) L8 AR-1262-2	9.127	7.888	40741456	18416184	2694.189	2401.381
38) L8 AR-1262-3	9.436	8.171	85140889	36486027	8149.918	11025.218m#
39) L8 AR-1262-4	9.532	8.238	85468996	39934341	11563.717	6954.995 #
40) L8 AR-1262-5	10.193	8.730	54759769	25165963	10559.340	9148.427

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012921\
Data File : P0075358.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Jan 2021 1:34
Operator : AJ/MA
Sample : M1246-05 10X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
250EM-007-0002

Manual Integrations
APPROVED

Sohil
2/1/2021 2:50:34 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 03:24:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 09:09:46 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

