

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012924\
 Data File : P0101452.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jan 2024 12:21
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 20:33:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 04:51:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.331	3.630	181.8E6	116.6E6	47.366	45.132
2)	SA Decachlor...	10.061	8.676	108.4E6	84983383	46.539	44.846
Target Compounds							
21)	L5 AR-1248-1	5.499	4.722	31002544	21965934	534.877	490.913
22)	L5 AR-1248-2	5.773	4.960	50392250	33496513	500.379	476.396
23)	L5 AR-1248-3	5.977	5.003	51731424	34571747	499.275	475.307
24)	L5 AR-1248-4	6.382	5.176	47141661	40151678	519.122	472.311
25)	L5 AR-1248-5	6.421	5.572	46551996	33091015	521.906	483.288



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012924\
Data File : PO101452.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2024 12:21
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 20:33:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 04:51:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.331 min
Delta R.T.: -0.003 min
Response: 181817219
Conc: 47.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.630 min
Delta R.T.: -0.001 min
Response: 116635727
Conc: 45.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.061 min
Delta R.T.: 0.003 min
Response: 108440824
Conc: 46.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.676 min
Delta R.T.: 0.001 min
Response: 84983383
Conc: 44.85 ng/ml



#21 AR-1248-1


R.T.: 5.499 min
Delta R.T.: -0.002 min
Response: 31002544
Conc: 534.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.722 min
Delta R.T.: -0.002 min
Response: 21965934
Conc: 490.91 ng/ml




#22 AR-1248-2

R.T.: 5.773 min
Delta R.T.: -0.002 min
Response: 50392250
Conc: 500.38 ng/ml



#22 AR-1248-2


R.T.: 4.960 min
Delta R.T.: -0.002 min
Response: 33496513
Conc: 476.40 ng/ml



#23 AR-1248-3


R.T.: 5.977 min
Delta R.T.: -0.002 min
Response: 51731424
Conc: 499.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.003 min
Delta R.T.: -0.002 min
Response: 34571747
Conc: 475.31 ng/ml




#24 AR-1248-4

R.T.: 6.382 min
Delta R.T.: -0.001 min
Response: 47141661
Conc: 519.12 ng/ml



#24 AR-1248-4


R.T.: 5.176 min
Delta R.T.: -0.001 min
Response: 40151678
Conc: 472.31 ng/ml



#25 AR-1248-5

R.T.: 6.421 min
Delta R.T.: -0.001 min
Response: 46551996
Conc: 521.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.572 min
Delta R.T.: -0.001 min
Response: 33091015
Conc: 483.29 ng/ml