

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0013020\
 Data File : P0066094.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Jan 2020 17:12
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 17:46:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 30 12:37:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.155	3.432	1325732	1905197	45.510	46.577
2)	SA Decachlor...	9.646	8.326	1596051	2414168	40.500	41.190
Target Compounds							
3)	L1 AR-1016-1	5.304	4.489	622420	765550	473.650	439.917
4)	L1 AR-1016-2	5.325	4.506	930826	1317663	474.023	477.819
5)	L1 AR-1016-3	5.385	4.679	562818	633656	475.458	455.161
6)	L1 AR-1016-4	5.483	4.721	470411	543365	475.643	470.315
7)	L1 AR-1016-5	5.773	4.930	475005	694438	469.526	465.605
31)	L7 AR-1260-1	6.884	5.947	978714	1518456	488.157	472.031
32)	L7 AR-1260-2	7.140	6.134	1353785	2173390	491.349	482.357
33)	L7 AR-1260-3	7.495	6.284	1137736	1756402	496.901	465.954
34)	L7 AR-1260-4	7.718	6.750	1078299	1496520	473.040	478.191
35)	L7 AR-1260-5	8.024	6.993	2260434	3997775	456.407	482.750



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO013020\
Data File : PO066094.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Jan 2020 17:12
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 17:46:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO013020.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 30 12:37:11 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.155 min
Delta R.T.: 0.000 min
Response: 1325732
Conc: 45.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.432 min
Delta R.T.: 0.002 min
Response: 1905197
Conc: 46.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.646 min
Delta R.T.: -0.002 min
Response: 1596051
Conc: 40.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.326 min
Delta R.T.: -0.001 min
Response: 2414168
Conc: 41.19 ng/ml



#3 AR-1016-1


R.T.: 5.304 min
Delta R.T.: 0.000 min
Response: 622420
Conc: 473.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.489 min
Delta R.T.: 0.001 min
Response: 765550
Conc: 439.92 ng/ml




#4 AR-1016-2

R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 930826
Conc: 474.02 ng/ml



#4 AR-1016-2


R.T.: 4.506 min
Delta R.T.: 0.000 min
Response: 1317663
Conc: 477.82 ng/ml



#5 AR-1016-3


R.T.: 5.385 min
Delta R.T.: 0.000 min
Response: 562818
Conc: 475.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.679 min
Delta R.T.: 0.001 min
Response: 633656
Conc: 455.16 ng/ml




#6 AR-1016-4

R.T.: 5.483 min
Delta R.T.: 0.000 min
Response: 470411
Conc: 475.64 ng/ml



#6 AR-1016-4


R.T.: 4.721 min
Delta R.T.: 0.000 min
Response: 543365
Conc: 470.32 ng/ml



#7 AR-1016-5


R.T.: 5.773 min
Delta R.T.: 0.000 min
Response: 475005
Conc: 469.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 4.930 min
Delta R.T.: 0.000 min
Response: 694438
Conc: 465.60 ng/ml




#31 AR-1260-1

R.T.: 6.884 min
Delta R.T.: -0.001 min
Response: 978714
Conc: 488.16 ng/ml



#31 AR-1260-1


R.T.: 5.947 min
Delta R.T.: 0.000 min
Response: 1518456
Conc: 472.03 ng/ml



#32 AR-1260-2


R.T.: 7.140 min
Delta R.T.: -0.001 min
Response: 1353785
Conc: 491.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.134 min
Delta R.T.: 0.000 min
Response: 2173390
Conc: 482.36 ng/ml




#33 AR-1260-3

R.T.: 7.495 min
Delta R.T.: 0.000 min
Response: 1137736
Conc: 496.90 ng/ml



#33 AR-1260-3


R.T.: 6.284 min
Delta R.T.: -0.001 min
Response: 1756402
Conc: 465.95 ng/ml



#34 AR-1260-4


R.T.: 7.718 min
Delta R.T.: -0.001 min
Response: 1078299
Conc: 473.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.750 min
Delta R.T.: 0.000 min
Response: 1496520
Conc: 478.19 ng/ml



#35 AR-1260-5

R.T.: 8.024 min
Delta R.T.: -0.001 min
Response: 2260434
Conc: 456.41 ng/ml



#35 AR-1260-5

R.T.: 6.993 min
Delta R.T.: 0.000 min
Response: 3997775
Conc: 482.75 ng/ml