

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020123\
 Data File : P0092333.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2023 21:01
 Operator : YP/AJ
 Sample : PB150607BS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB150607BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 00:50:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 04:40:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.447	3.645	45154749	24226677	15.598	17.254
2)	SA Decachlor...	10.364	8.686	43197473	22590903	20.930	20.011
Target Compounds							
3)	L1 AR-1016-1	5.633	4.733	37724196	17687842	412.478	384.847
4)	L1 AR-1016-2	5.655	4.751	55278984	25260883	412.808	385.205
5)	L1 AR-1016-3	5.718	4.928	35621802	13313797	413.945	376.576
6)	L1 AR-1016-4	5.818	4.970	27800515	12086594	417.013	388.549
7)	L1 AR-1016-5	6.119	5.184	29418840	15051314	409.446	378.497
31)	L7 AR-1260-1	7.269	6.222	54744215	28122306	415.498	372.286
32)	L7 AR-1260-2	7.534	6.411	60142511	32935697	405.583	375.919
33)	L7 AR-1260-3	7.901	6.565	40733015	30022275	359.630	346.321
34)	L7 AR-1260-4	8.131	7.040	47850183	21771385	385.997	325.103
35)	L7 AR-1260-5	8.466	7.283	82257925	48767631	370.436	342.709



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020123\
Data File : P0092333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2023 21:01
Operator : YP/AJ
Sample : PB150607BS
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB150607BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 00:50:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 04:40:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.447 min
Delta R.T.: -0.002 min
Response: 45154749
Conc: 15.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB150607BS




#1 Tetrachloro-m-xylene

R.T.: 3.645 min
Delta R.T.: -0.001 min
Response: 24226677
Conc: 17.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.364 min
Delta R.T.: -0.007 min
Response: 43197473
Conc: 20.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.686 min
Delta R.T.: -0.007 min
Response: 22590903
Conc: 20.01 ng/ml



#3 AR-1016-1


R.T.: 5.633 min
Delta R.T.: -0.002 min
Response: 37724196
Conc: 412.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB150607BS




#3 AR-1016-1

R.T.: 4.733 min
Delta R.T.: -0.003 min
Response: 17687842
Conc: 384.85 ng/ml




#4 AR-1016-2

R.T.: 5.655 min
Delta R.T.: -0.003 min
Response: 55278984
Conc: 412.81 ng/ml



#4 AR-1016-2


R.T.: 4.751 min
Delta R.T.: -0.002 min
Response: 25260883
Conc: 385.20 ng/ml



#5 AR-1016-3


R.T.: 5.718 min
Delta R.T.: -0.003 min
Response: 35621802
Conc: 413.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB150607BS




#5 AR-1016-3

R.T.: 4.928 min
Delta R.T.: -0.003 min
Response: 13313797
Conc: 376.58 ng/ml




#6 AR-1016-4

R.T.: 5.818 min
Delta R.T.: -0.003 min
Response: 27800515
Conc: 417.01 ng/ml



#6 AR-1016-4


R.T.: 4.970 min
Delta R.T.: -0.003 min
Response: 12086594
Conc: 388.55 ng/ml



#7 AR-1016-5


R.T.: 6.119 min
Delta R.T.: -0.002 min
Response: 29418840
Conc: 409.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB150607BS




#7 AR-1016-5

R.T.: 5.184 min
Delta R.T.: -0.003 min
Response: 15051314
Conc: 378.50 ng/ml




#31 AR-1260-1

R.T.: 7.269 min
Delta R.T.: -0.003 min
Response: 54744215
Conc: 415.50 ng/ml



#31 AR-1260-1


R.T.: 6.222 min
Delta R.T.: -0.004 min
Response: 28122306
Conc: 372.29 ng/ml



#32 AR-1260-2


R.T.: 7.534 min
Delta R.T.: -0.003 min
Response: 60142511
Conc: 405.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB150607BS




#32 AR-1260-2

R.T.: 6.411 min
Delta R.T.: -0.004 min
Response: 32935697
Conc: 375.92 ng/ml




#33 AR-1260-3

R.T.: 7.901 min
Delta R.T.: -0.003 min
Response: 40733015
Conc: 359.63 ng/ml



#33 AR-1260-3


R.T.: 6.565 min
Delta R.T.: -0.005 min
Response: 30022275
Conc: 346.32 ng/ml



#34 AR-1260-4


R.T.: 8.131 min
Delta R.T.: -0.004 min
Response: 47850183
Conc: 386.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB150607BS




#34 AR-1260-4

R.T.: 7.040 min
Delta R.T.: -0.006 min
Response: 21771385
Conc: 325.10 ng/ml



#35 AR-1260-5

R.T.: 8.466 min
Delta R.T.: -0.003 min
Response: 82257925
Conc: 370.44 ng/ml



#35 AR-1260-5

R.T.: 7.283 min
Delta R.T.: -0.005 min
Response: 48767631
Conc: 342.71 ng/ml