

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020124\
 Data File : P0101507.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2024 10:50
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 20:46:42 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 04:51:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.330	3.629	184.8E6	120.2E6	48.139	46.514
2) SA Decachlor...	10.056	8.672	114.8E6	86159191	49.275	45.467
Target Compounds						
3) L1 AR-1016-1	5.497	4.721	52148708	37018904	533.471	500.748
4) L1 AR-1016-2	5.519	4.740	75116097	50461981	546.077	511.979
5) L1 AR-1016-3	5.581	4.918	49367578	27338669	550.309	498.322
6) L1 AR-1016-4	5.680	4.959	39113920	23915306	547.261	470.604
7) L1 AR-1016-5	5.975	5.175	39392661	30840494	498.353	472.392
31) L7 AR-1260-1	7.105	6.214	67184050	57218342	471.094	463.703
32) L7 AR-1260-2	7.363	6.404	74839773	67077904	510.910	490.582
33) L7 AR-1260-3	7.724	6.558	55811304	61393957	477.974	478.577
34) L7 AR-1260-4	7.951	7.032	59746660	50183696	507.496	485.571
35) L7 AR-1260-5	8.267	7.276	114.2E6	108.6E6	520.742	517.154


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020124\
Data File : PO101507.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2024 10:50
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 20:46:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 04:51:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.330 min
Delta R.T.: -0.004 min
Response: 184784845
Conc: 48.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.629 min
Delta R.T.: -0.002 min
Response: 120205950
Conc: 46.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.056 min
Delta R.T.: -0.002 min
Response: 114815659
Conc: 49.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.672 min
Delta R.T.: -0.003 min
Response: 86159191
Conc: 45.47 ng/ml



#3 AR-1016-1


R.T.: 5.497 min
Delta R.T.: -0.003 min
Response: 52148708
Conc: 533.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.721 min
Delta R.T.: -0.003 min
Response: 37018904
Conc: 500.75 ng/ml




#4 AR-1016-2

R.T.: 5.519 min
Delta R.T.: -0.003 min
Response: 75116097
Conc: 546.08 ng/ml



#4 AR-1016-2


R.T.: 4.740 min
Delta R.T.: -0.003 min
Response: 50461981
Conc: 511.98 ng/ml



#5 AR-1016-3


R.T.: 5.581 min
Delta R.T.: -0.003 min
Response: 49367578
Conc: 550.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.918 min
Delta R.T.: -0.003 min
Response: 27338669
Conc: 498.32 ng/ml




#6 AR-1016-4

R.T.: 5.680 min
Delta R.T.: -0.003 min
Response: 39113920
Conc: 547.26 ng/ml



#6 AR-1016-4


R.T.: 4.959 min
Delta R.T.: -0.003 min
Response: 23915306
Conc: 470.60 ng/ml



#7 AR-1016-5


R.T.: 5.975 min
Delta R.T.: -0.003 min
Response: 39392661
Conc: 498.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.175 min
Delta R.T.: -0.003 min
Response: 30840494
Conc: 472.39 ng/ml




#31 AR-1260-1

R.T.: 7.105 min
Delta R.T.: -0.003 min
Response: 67184050
Conc: 471.09 ng/ml



#31 AR-1260-1


R.T.: 6.214 min
Delta R.T.: -0.003 min
Response: 57218342
Conc: 463.70 ng/ml



#32 AR-1260-2


R.T.: 7.363 min
Delta R.T.: -0.002 min
Response: 74839773
Conc: 510.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.404 min
Delta R.T.: -0.002 min
Response: 67077904
Conc: 490.58 ng/ml




#33 AR-1260-3

R.T.: 7.724 min
Delta R.T.: -0.002 min
Response: 55811304
Conc: 477.97 ng/ml



#33 AR-1260-3


R.T.: 6.558 min
Delta R.T.: -0.003 min
Response: 61393957
Conc: 478.58 ng/ml



#34 AR-1260-4


R.T.: 7.951 min
Delta R.T.: -0.001 min
Response: 59746660
Conc: 507.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.032 min
Delta R.T.: -0.003 min
Response: 50183696
Conc: 485.57 ng/ml



#35 AR-1260-5

R.T.: 8.267 min
Delta R.T.: 0.000 min
Response: 114192797
Conc: 520.74 ng/ml



#35 AR-1260-5

R.T.: 7.276 min
Delta R.T.: -0.003 min
Response: 108555550
Conc: 517.15 ng/ml