

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020124\
 Data File : P0101540.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2024 21:17
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 03:53:32 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 04:51:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.330	3.629	176.3E6	115.2E6	45.931	44.586
2) SA Decachlor...	10.058	8.673	110.3E6	86170302	47.319	45.472
Target Compounds						
26) L6 AR-1254-1	6.355	5.529	53564175	58787811	450.196	472.527
27) L6 AR-1254-2	6.573	5.677	79568044	52550883	467.168	469.467
28) L6 AR-1254-3	6.941	6.085	82533661	79289362	502.572	476.518
29) L6 AR-1254-4	7.227	6.315	53524991	44653413	535.504	523.507
30) L6 AR-1254-5	7.648	6.735	60081977	65479957	511.204	483.254



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020124\
Data File : PO101540.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2024 21:17
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 03:53:32 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 04:51:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.330 min
Delta R.T.: -0.004 min
Response: 176305881
Conc: 45.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.629 min
Delta R.T.: -0.002 min
Response: 115224425
Conc: 44.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.058 min
Delta R.T.: 0.000 min
Response: 110258886
Conc: 47.32 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.673 min
Delta R.T.: -0.002 min
Response: 86170302
Conc: 45.47 ng/ml



#26 AR-1254-1


R.T.: 6.355 min
Delta R.T.: -0.003 min
Response: 53564175
Conc: 450.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.529 min
Delta R.T.: -0.003 min
Response: 58787811
Conc: 472.53 ng/ml




#27 AR-1254-2

R.T.: 6.573 min
Delta R.T.: -0.003 min
Response: 79568044
Conc: 467.17 ng/ml



#27 AR-1254-2


R.T.: 5.677 min
Delta R.T.: -0.003 min
Response: 52550883
Conc: 469.47 ng/ml



#28 AR-1254-3


R.T.: 6.941 min
Delta R.T.: -0.003 min
Response: 82533661
Conc: 502.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.085 min
Delta R.T.: -0.004 min
Response: 79289362
Conc: 476.52 ng/ml




#29 AR-1254-4

R.T.: 7.227 min
Delta R.T.: -0.003 min
Response: 53524991
Conc: 535.50 ng/ml



#29 AR-1254-4


R.T.: 6.315 min
Delta R.T.: -0.003 min
Response: 44653413
Conc: 523.51 ng/ml



#30 AR-1254-5

R.T.: 7.648 min
Delta R.T.: -0.002 min
Response: 60081977
Conc: 511.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.735 min
Delta R.T.: -0.003 min
Response: 65479957
Conc: 483.25 ng/ml