

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020124\
 Data File : P0101569.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Feb 2024 06:27
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 06:49:43 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 04:51:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.330	3.628	172.6E6	115.2E6	44.973	44.562
2) SA Decachlor...	10.056	8.672	104.8E6	82429318	44.978	43.498
Target Compounds						
21) L5 AR-1248-1	5.497	4.720	29502417	22875270	508.996	511.235
22) L5 AR-1248-2	5.771	4.958	46964967	32570614	466.347	463.228
23) L5 AR-1248-3	5.975	5.000	48802124	34756288	471.003	477.845
24) L5 AR-1248-4	6.380	5.174	47343216	40342104	521.342	474.551
25) L5 AR-1248-5	6.419	5.570	47381187	35871762	531.202	523.901



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020124\
Data File : PO101569.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Feb 2024 06:27
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 06:49:43 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 04:51:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.330 min
Delta R.T.: -0.004 min
Response: 172629271
Conc: 44.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.628 min
Delta R.T.: -0.003 min
Response: 115163593
Conc: 44.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.056 min
Delta R.T.: -0.002 min
Response: 104803232
Conc: 44.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.672 min
Delta R.T.: -0.003 min
Response: 82429318
Conc: 43.50 ng/ml



#21 AR-1248-1


R.T.: 5.497 min
Delta R.T.: -0.003 min
Response: 29502417
Conc: 509.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.720 min
Delta R.T.: -0.004 min
Response: 22875270
Conc: 511.24 ng/ml




#22 AR-1248-2

R.T.: 5.771 min
Delta R.T.: -0.004 min
Response: 46964967
Conc: 466.35 ng/ml



#22 AR-1248-2


R.T.: 4.958 min
Delta R.T.: -0.004 min
Response: 32570614
Conc: 463.23 ng/ml



#23 AR-1248-3


R.T.: 5.975 min
Delta R.T.: -0.004 min
Response: 48802124
Conc: 471.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.000 min
Delta R.T.: -0.004 min
Response: 34756288
Conc: 477.84 ng/ml




#24 AR-1248-4

R.T.: 6.380 min
Delta R.T.: -0.004 min
Response: 47343216
Conc: 521.34 ng/ml



#24 AR-1248-4


R.T.: 5.174 min
Delta R.T.: -0.004 min
Response: 40342104
Conc: 474.55 ng/ml



#25 AR-1248-5

R.T.: 6.419 min
Delta R.T.: -0.003 min
Response: 47381187
Conc: 531.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.570 min
Delta R.T.: -0.003 min
Response: 35871762
Conc: 523.90 ng/ml