

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020223\
 Data File : P0092369.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Feb 2023 17:00
 Operator : YP/AJ
 Sample : PB150630BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 17:39:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 04:40:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.447	3.644	49999952	21966904	17.272	15.644
2)	SA Decachlor...	10.372	8.687	39609189	20858456	19.191	18.477
Target Compounds							
8)	L2 AR-1221-1	0.000	3.944f	0	315572	N.D.	16.670 #
9)	L2 AR-1221-2	0.000	3.944	0	315572	N.D.	22.454 #
20)	L4 AR-1242-5	0.000	5.577f	0	437162	N.D.	12.076 #
25)	L5 AR-1248-5	0.000	5.577	0	437162	N.D.	9.234 #
26)	L6 AR-1254-1	0.000	5.577f	0	437162	N.D.	5.577 #
33)	L7 AR-1260-3	0.000	6.559	0	490838	N.D.	5.662 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020223\
Data File : PO092369.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Feb 2023 17:00
Operator : YP/AJ
Sample : PB150630BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 17:39:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 04:40:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.447 min
Delta R.T.: -0.002 min
Response: 49999952
Conc: 17.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: -0.002 min
Response: 21966904
Conc: 15.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.372 min
Delta R.T.: 0.000 min
Response: 39609189
Conc: 19.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.687 min
Delta R.T.: -0.006 min
Response: 20858456
Conc: 18.48 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 4.659 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 3.944 min
Delta R.T.: 0.067 min
Response: 315572
Conc: 16.67 ng/ml




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T.: 4.746 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 3.944 min
Delta R.T.: -0.018 min
Response: 315572
Conc: 22.45 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 6.571 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#20 AR-1242-5

R.T.: 5.577 min
Delta R.T.: 0.039 min
Response: 437162
Conc: 12.08 ng/ml




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T.: 6.573 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 5.577 min
Delta R.T.: -0.004 min
Response: 437162
Conc: 9.23 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T.: 6.506 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 5.577 min
Delta R.T.: 0.038 min
Response: 437162
Conc: 5.58 ng/ml



#33 AR-1260-3

R.T.: 0.000 min
Exp R.T.: 7.905 min
Response: 0
Conc: N.D.



#33 AR-1260-3

R.T.: 6.559 min
Delta R.T.: -0.011 min
Response: 490838
Conc: 5.66 ng/ml