

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020224\
 Data File : P0101573.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Feb 2024 09:19
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 11:42:01 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 04:51:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.329	3.627	192.3E6	127.0E6	50.108	49.132
2) SA Decachlor...	10.056	8.672	116.3E6	92247363	49.902	48.679
Target Compounds						
3) L1 AR-1016-1	5.496	4.719	53660433	38761956	548.936	524.326
4) L1 AR-1016-2	5.519	4.738	77311353	53910686	562.036	546.969
5) L1 AR-1016-3	5.581	4.917	50616375	28546352	564.230	520.336
6) L1 AR-1016-4	5.679	4.958	39891919	24760428	558.147	487.234
7) L1 AR-1016-5	5.975	5.174	39918014	32327049	505.000	495.162
31) L7 AR-1260-1	7.105	6.213	70352482	60960637	493.311	494.031
32) L7 AR-1260-2	7.363	6.403	76526949	70869606	522.427	518.313
33) L7 AR-1260-3	7.724	6.558	58414419	64512349	500.267	502.885
34) L7 AR-1260-4	7.951	7.032	62473278	52497211	530.656	507.956
35) L7 AR-1260-5	8.266	7.276	117.5E6	113.6E6	535.915	541.182



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020224\
Data File : PO101573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Feb 2024 09:19
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 11:42:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 04:51:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.329 min
Delta R.T.: -0.005 min
Response: 192341983
Conc: 50.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.627 min
Delta R.T.: -0.004 min
Response: 126974034
Conc: 49.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.056 min
Delta R.T.: -0.002 min
Response: 116277509
Conc: 49.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.672 min
Delta R.T.: -0.003 min
Response: 92247363
Conc: 48.68 ng/ml



#3 AR-1016-1


R.T.: 5.496 min
Delta R.T.: -0.004 min
Response: 53660433
Conc: 548.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.719 min
Delta R.T.: -0.004 min
Response: 38761956
Conc: 524.33 ng/ml




#4 AR-1016-2

R.T.: 5.519 min
Delta R.T.: -0.003 min
Response: 77311353
Conc: 562.04 ng/ml



#4 AR-1016-2


R.T.: 4.738 min
Delta R.T.: -0.004 min
Response: 53910686
Conc: 546.97 ng/ml



#5 AR-1016-3


R.T.: 5.581 min
Delta R.T.: -0.004 min
Response: 50616375
Conc: 564.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.917 min
Delta R.T.: -0.004 min
Response: 28546352
Conc: 520.34 ng/ml




#6 AR-1016-4

R.T.: 5.679 min
Delta R.T.: -0.004 min
Response: 39891919
Conc: 558.15 ng/ml



#6 AR-1016-4


R.T.: 4.958 min
Delta R.T.: -0.004 min
Response: 24760428
Conc: 487.23 ng/ml



#7 AR-1016-5


R.T.: 5.975 min
Delta R.T.: -0.004 min
Response: 39918014
Conc: 505.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.174 min
Delta R.T.: -0.004 min
Response: 32327049
Conc: 495.16 ng/ml




#31 AR-1260-1

R.T.: 7.105 min
Delta R.T.: -0.003 min
Response: 70352482
Conc: 493.31 ng/ml



#31 AR-1260-1


R.T.: 6.213 min
Delta R.T.: -0.004 min
Response: 60960637
Conc: 494.03 ng/ml



#32 AR-1260-2


R.T.: 7.363 min
Delta R.T.: -0.002 min
Response: 76526949
Conc: 522.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.403 min
Delta R.T.: -0.003 min
Response: 70869606
Conc: 518.31 ng/ml




#33 AR-1260-3

R.T.: 7.724 min
Delta R.T.: -0.003 min
Response: 58414419
Conc: 500.27 ng/ml



#33 AR-1260-3


R.T.: 6.558 min
Delta R.T.: -0.003 min
Response: 64512349
Conc: 502.89 ng/ml



#34 AR-1260-4


R.T.: 7.951 min
Delta R.T.: -0.002 min
Response: 62473278
Conc: 530.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.032 min
Delta R.T.: -0.003 min
Response: 52497211
Conc: 507.96 ng/ml



#35 AR-1260-5

R.T.: 8.266 min
Delta R.T.: -0.002 min
Response: 117520106
Conc: 535.91 ng/ml



#35 AR-1260-5

R.T.: 7.276 min
Delta R.T.: -0.003 min
Response: 113599145
Conc: 541.18 ng/ml