

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020224\
 Data File : P0101582.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Feb 2024 12:10
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 01:30:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 04:51:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.330	3.629	183.1E6	125.7E6	47.709	48.657
2)	SA Decachlor...	10.054	8.672	109.9E6	85371281	47.145	45.051
Target Compounds							
16)	L4 AR-1242-1	5.497	4.720	38639598	31113487	499.380	528.825
17)	L4 AR-1242-2	5.519	4.739	54573696	42788300	503.933	551.790
18)	L4 AR-1242-3	5.581	4.917	34004946	22858162	478.402	527.304
19)	L4 AR-1242-4	5.679	5.000	27056209	24889614	475.035	500.045
20)	L4 AR-1242-5	6.418	5.528	31146460	27505019	580.399	519.228



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020224\
Data File : PO101582.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Feb 2024 12:10
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 01:30:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 04:51:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.330 min
Delta R.T.: -0.004 min
Response: 183133691
Conc: 47.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.629 min
Delta R.T.: -0.003 min
Response: 125745448
Conc: 48.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.054 min
Delta R.T.: -0.003 min
Response: 109852377
Conc: 47.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.672 min
Delta R.T.: -0.003 min
Response: 85371281
Conc: 45.05 ng/ml



#16 AR-1242-1


R.T.: 5.497 min
Delta R.T.: -0.003 min
Response: 38639598
Conc: 499.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.720 min
Delta R.T.: -0.003 min
Response: 31113487
Conc: 528.82 ng/ml




#17 AR-1242-2

R.T.: 5.519 min
Delta R.T.: -0.003 min
Response: 54573696
Conc: 503.93 ng/ml



#17 AR-1242-2


R.T.: 4.739 min
Delta R.T.: -0.004 min
Response: 42788300
Conc: 551.79 ng/ml



#18 AR-1242-3


R.T.: 5.581 min
Delta R.T.: -0.003 min
Response: 34004946
Conc: 478.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.917 min
Delta R.T.: -0.004 min
Response: 22858162
Conc: 527.30 ng/ml




#19 AR-1242-4

R.T.: 5.679 min
Delta R.T.: -0.004 min
Response: 27056209
Conc: 475.03 ng/ml



#19 AR-1242-4


R.T.: 5.000 min
Delta R.T.: -0.004 min
Response: 24889614
Conc: 500.04 ng/ml



#20 AR-1242-5

R.T.: 6.418 min
Delta R.T.: -0.002 min
Response: 31146460
Conc: 580.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.528 min
Delta R.T.: -0.003 min
Response: 27505019
Conc: 519.23 ng/ml