

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020224\
 Data File : P0101614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Feb 2024 22:30
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 02:05:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 04:51:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.328	3.628	181.4E6	122.0E6	47.251	47.200
2) SA Decachlor...	10.052	8.671	114.5E6	88623617	49.143	46.767
Target Compounds						
26) L6 AR-1254-1	6.352	5.528	56374922	60991531	473.820	490.240
27) L6 AR-1254-2	6.571	5.676	82556150	53873344	484.712	481.282
28) L6 AR-1254-3	6.939	6.084	82667042	82826572	503.384	497.776
29) L6 AR-1254-4	7.225	6.314	55722903	48199536	557.494	565.081
30) L6 AR-1254-5	7.646	6.733	61339814	70173423	521.906	517.893



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020224\
Data File : PO101614.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Feb 2024 22:30
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 02:05:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 04:51:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.328 min
Delta R.T.: -0.006 min
Response: 181373485
Conc: 47.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.628 min
Delta R.T.: -0.003 min
Response: 121980831
Conc: 47.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.052 min
Delta R.T.: -0.005 min
Response: 114508773
Conc: 49.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.671 min
Delta R.T.: -0.004 min
Response: 88623617
Conc: 46.77 ng/ml



#26 AR-1254-1


R.T.: 6.352 min
Delta R.T.: -0.005 min
Response: 56374922
Conc: 473.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.528 min
Delta R.T.: -0.004 min
Response: 60991531
Conc: 490.24 ng/ml




#27 AR-1254-2

R.T.: 6.571 min
Delta R.T.: -0.005 min
Response: 82556150
Conc: 484.71 ng/ml



#27 AR-1254-2


R.T.: 5.676 min
Delta R.T.: -0.004 min
Response: 53873344
Conc: 481.28 ng/ml



#28 AR-1254-3


R.T.: 6.939 min
Delta R.T.: -0.005 min
Response: 82667042
Conc: 503.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.084 min
Delta R.T.: -0.005 min
Response: 82826572
Conc: 497.78 ng/ml




#29 AR-1254-4

R.T.: 7.225 min
Delta R.T.: -0.004 min
Response: 55722903
Conc: 557.49 ng/ml



#29 AR-1254-4


R.T.: 6.314 min
Delta R.T.: -0.004 min
Response: 48199536
Conc: 565.08 ng/ml



#30 AR-1254-5

R.T.: 7.646 min
Delta R.T.: -0.004 min
Response: 61339814
Conc: 521.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.733 min
Delta R.T.: -0.005 min
Response: 70173423
Conc: 517.89 ng/ml