

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020224\
 Data File : P0101626.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2024 02:24
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 05:38:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 04:51:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.331	3.630	173.8E6	118.5E6	45.269	45.865
2) SA Decachlor...	10.058	8.674	109.5E6	85243125	46.973	44.983
Target Compounds						
21) L5 AR-1248-1	5.498	4.722	30170677	23755532	520.525	530.908
22) L5 AR-1248-2	5.773	4.960	47918737	33892105	475.818	482.023
23) L5 AR-1248-3	5.976	5.002	50120176	35593250	483.724	489.352
24) L5 AR-1248-4	6.381	5.176	51016914	41327089	561.796	486.137
25) L5 AR-1248-5	6.419	5.572	50470138	37556743	565.833	548.510



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020224\
Data File : PO101626.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2024 02:24
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 05:38:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 04:51:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.331 min
Delta R.T.: -0.003 min
Response: 173767559
Conc: 45.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.630 min
Delta R.T.: 0.000 min
Response: 118529869
Conc: 45.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.058 min
Delta R.T.: 0.000 min
Response: 109451939
Conc: 46.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.674 min
Delta R.T.: 0.000 min
Response: 85243125
Conc: 44.98 ng/ml



#21 AR-1248-1


R.T.: 5.498 min
Delta R.T.: -0.003 min
Response: 30170677
Conc: 520.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.722 min
Delta R.T.: -0.002 min
Response: 23755532
Conc: 530.91 ng/ml




#22 AR-1248-2

R.T.: 5.773 min
Delta R.T.: -0.003 min
Response: 47918737
Conc: 475.82 ng/ml



#22 AR-1248-2


R.T.: 4.960 min
Delta R.T.: -0.002 min
Response: 33892105
Conc: 482.02 ng/ml



#23 AR-1248-3


R.T.: 5.976 min
Delta R.T.: -0.003 min
Response: 50120176
Conc: 483.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.002 min
Delta R.T.: -0.002 min
Response: 35593250
Conc: 489.35 ng/ml




#24 AR-1248-4

R.T.: 6.381 min
Delta R.T.: -0.002 min
Response: 51016914
Conc: 561.80 ng/ml



#24 AR-1248-4


R.T.: 5.176 min
Delta R.T.: -0.002 min
Response: 41327089
Conc: 486.14 ng/ml



#25 AR-1248-5

R.T.: 6.419 min
Delta R.T.: -0.003 min
Response: 50470138
Conc: 565.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.572 min
Delta R.T.: -0.002 min
Response: 37556743
Conc: 548.51 ng/ml