

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020321\
 Data File : P0075442.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Feb 2021 15:59
 Operator : AJ/MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 16:51:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 09:09:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.924	3.996	4548750	2341398	44.947	48.707
2) SA Decachlor...	10.946	9.248	4684397	2113240	48.552	43.535
Target Compounds						
3) L1 AR-1016-1	6.249	5.240	1674593	894319	423.837	460.758
4) L1 AR-1016-2	6.273	5.260	2332165	1242459	425.238	463.296
5) L1 AR-1016-3	6.339	5.449	1419762	671262	434.085	467.582
6) L1 AR-1016-4	6.446	5.502	1172770	557254	425.917	464.778
7) L1 AR-1016-5	6.763	5.729	1121966	683677	420.955	469.128
31) L7 AR-1260-1	7.944	6.815	2062738	1144254	447.111	421.498
32) L7 AR-1260-2	8.210	7.013	2892154	1407289	448.083	424.219
33) L7 AR-1260-3	8.576	7.165	2474258	1252518	453.430	421.437
34) L7 AR-1260-4	8.807	7.645	2385639	1080135	464.087	415.208
35) L7 AR-1260-5	9.135	7.892	5711536	2654453	483.978	419.455



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020321\
Data File : PO075442.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Feb 2021 15:59
Operator : AJ/MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 16:51:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 09:09:46 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.924 min
Delta R.T.: 0.000 min
Response: 4548750
Conc: 44.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.996 min
Delta R.T.: 0.000 min
Response: 2341398
Conc: 48.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.946 min
Delta R.T.: 0.008 min
Response: 4684397
Conc: 48.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.248 min
Delta R.T.: 0.002 min
Response: 2113240
Conc: 43.53 ng/ml



#3 AR-1016-1


R.T.: 6.249 min
Delta R.T.: 0.002 min
Response: 1674593
Conc: 423.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.240 min
Delta R.T.: 0.000 min
Response: 894319
Conc: 460.76 ng/ml




#4 AR-1016-2

R.T.: 6.273 min
Delta R.T.: 0.002 min
Response: 2332165
Conc: 425.24 ng/ml



#4 AR-1016-2


R.T.: 5.260 min
Delta R.T.: 0.000 min
Response: 1242459
Conc: 463.30 ng/ml



#5 AR-1016-3


R.T.: 6.339 min
Delta R.T.: 0.002 min
Response: 1419762
Conc: 434.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.449 min
Delta R.T.: 0.000 min
Response: 671262
Conc: 467.58 ng/ml




#6 AR-1016-4

R.T.: 6.446 min
Delta R.T.: 0.001 min
Response: 1172770
Conc: 425.92 ng/ml



#6 AR-1016-4


R.T.: 5.502 min
Delta R.T.: 0.000 min
Response: 557254
Conc: 464.78 ng/ml



#7 AR-1016-5


R.T.: 6.763 min
Delta R.T.: 0.002 min
Response: 1121966
Conc: 420.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.729 min
Delta R.T.: 0.000 min
Response: 683677
Conc: 469.13 ng/ml




#31 AR-1260-1

R.T.: 7.944 min
Delta R.T.: 0.004 min
Response: 2062738
Conc: 447.11 ng/ml



#31 AR-1260-1


R.T.: 6.815 min
Delta R.T.: 0.000 min
Response: 1144254
Conc: 421.50 ng/ml



#32 AR-1260-2


R.T.: 8.210 min
Delta R.T.: 0.003 min
Response: 2892154
Conc: 448.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.013 min
Delta R.T.: 0.000 min
Response: 1407289
Conc: 424.22 ng/ml




#33 AR-1260-3

R.T.: 8.576 min
Delta R.T.: 0.004 min
Response: 2474258
Conc: 453.43 ng/ml



#33 AR-1260-3


R.T.: 7.165 min
Delta R.T.: 0.000 min
Response: 1252518
Conc: 421.44 ng/ml



#34 AR-1260-4


R.T.: 8.807 min
Delta R.T.: 0.003 min
Response: 2385639
Conc: 464.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.645 min
Delta R.T.: 0.000 min
Response: 1080135
Conc: 415.21 ng/ml



#35 AR-1260-5

R.T.: 9.135 min
Delta R.T.: 0.005 min
Response: 5711536
Conc: 483.98 ng/ml



#35 AR-1260-5

R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 2654453
Conc: 419.45 ng/ml