

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020325\
 Data File : P0109390.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2025 00:22
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 00:57:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0020325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 04 00:51:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.698	3.695	197.0E6	135.9E6	25.396	25.278
2) SA Decachlor...	8.758	8.709	158.8E6	83098594	25.997	26.232
Target Compounds						
26) L6 AR-1254-1	5.602	5.581	108.9E6	72866460	264.613	264.885
27) L6 AR-1254-2	5.751	5.728	96568808	65137257	266.238	268.775
28) L6 AR-1254-3	6.157	6.132	153.5E6	100.5E6	260.492	260.917
29) L6 AR-1254-4	6.386	6.359	87468245	55136004	258.317	259.656
30) L6 AR-1254-5	6.808	6.777	133.2E6	83976710	259.254	260.917

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020325\
Data File : PO109390.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2025 00:22
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 00:57:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO020325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 04 00:51:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

