

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020420\
 Data File : P0066259.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Feb 2020 21:55
 Operator : DD\AJ
 Sample : PB126560BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB126560BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 04:46:19 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 31 01:14:46 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.154	3.432	748554	976390	25.696	23.870
2)	SA Decachlor...	9.646	8.327	640039	887301	16.241	15.139
Target Compounds							
3)	L1 AR-1016-1	5.303	4.488	370293	367559	281.786	211.215 #
4)	L1 AR-1016-2	5.325	4.506	473905	683989	241.336	248.032
5)	L1 AR-1016-3	5.386	4.678	303619	315451	256.492	226.591
6)	L1 AR-1016-4	5.483	4.720	582095	230789	588.570	199.762 #
7)	L1 AR-1016-5	5.772	4.929	376750	306683	372.404	205.624 #
8)	L2 AR-1221-1	4.359	3.641	23960	25442	67.276	50.315 #
9)	L2 AR-1221-2	4.448	3.724	38195	104531	145.437	280.736 #
10)	L2 AR-1221-3	4.522	3.796	159665	265298	182.623	221.850
11)	L3 AR-1232-1	4.522	3.796	159665	265298	235.754	285.583
12)	L3 AR-1232-2	5.040	4.506	195616	683989	469.335	617.406 #
13)	L3 AR-1232-3	5.325	4.678	473905	315451	581.633	565.828
14)	L3 AR-1232-4	5.483	4.761	582095	268557	1422.089	556.451 #
15)	L3 AR-1232-5	5.573	4.929	547788	306683	2000.420	594.617 #
16)	L4 AR-1242-1	5.303	4.488	370293	367559	365.611	277.103
17)	L4 AR-1242-2	5.325	4.506	473905	683989	319.695	322.573
18)	L4 AR-1242-3	5.386	4.678	303619	315451	333.181	296.784
19)	L4 AR-1242-4	5.483	4.761	582095	268557	762.821	261.257 #
20)	L4 AR-1242-5	6.210	5.275	32125	249510	34.638	180.439 #
21)	L5 AR-1248-1	5.303	4.488	370293	367559	485.723	353.809 #
22)	L5 AR-1248-2	5.573	4.720	547788	230789	505.211	155.344 #
23)	L5 AR-1248-3	5.772	4.761	376750	268557	308.537	179.611 #
24)	L5 AR-1248-4	6.171	4.929	38788	306683	24.575	169.878 #
25)	L5 AR-1248-5	6.210	5.315	32125	35916	21.610	18.366
26)	L6 AR-1254-1	6.144	5.275	180785	249510	116.357	78.386 #
27)	L6 AR-1254-2	6.360	5.421	187476	257129	75.448	87.615
28)	L6 AR-1254-3	6.723	5.833	101308	478063	37.913	104.102 #
29)	L6 AR-1254-4	7.005	6.046	63400	53208	27.457	16.887 #
30)	L6 AR-1254-5	7.420	6.458	667591	994460	292.578	234.794
31)	L7 AR-1260-1	6.884	5.947	443340	671464	221.127	208.733
32)	L7 AR-1260-2	7.140	6.135	620089	969145	225.058	215.090
33)	L7 AR-1260-3	7.495	6.285	417524	763883	182.352	202.649
34)	L7 AR-1260-4	7.719	6.751	438821	601552	192.507	192.217
35)	L7 AR-1260-5	8.024	6.994	861116	1593046	173.869	192.368
36)	L8 AR-1262-1	7.495	6.548	417524	408578	126.625	193.936 #
37)	L8 AR-1262-2	8.024	6.994	861116	1593046	157.984	168.340
38)	L8 AR-1262-3	8.323	7.272	547410	819791	152.122	213.011 #
39)	L8 AR-1262-4	8.370	7.336	281108	2588879	103.302	393.923 #
40)	L8 AR-1262-5	8.984	7.829	166600	262390	87.667	93.029
41)	L9 AR-1268-1	8.323	7.272	547410	819791	79.326	68.271

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020420\
Data File : P0066259.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Feb 2020 21:55
Operator : DD\AJ
Sample : PB126560BS
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB126560BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 04:46:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 31 01:14:46 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.370	7.336	281108	2588879	46.666	250.675 #
43)	L9 AR-1268-3	8.596	7.537	111628	68930	21.506	7.911 #
44)	L9 AR-1268-4	8.984	7.829	166600	262390	78.782	79.441
45)	L9 AR-1268-5	9.352	8.098	60066	92737	3.951	3.743



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020420\
Data File : P0066259.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Feb 2020 21:55
Operator : DD\AJ
Sample : PB126560BS
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
PB126560BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 04:46:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 31 01:14:46 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.154 min
Delta R.T.: -0.011 min
Response: 748554
Conc: 25.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#1 Tetrachloro-m-xylene

R.T.: 3.432 min
Delta R.T.: -0.001 min
Response: 976390
Conc: 23.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.646 min
Delta R.T.: -0.016 min
Response: 640039
Conc: 16.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.327 min
Delta R.T.: -0.005 min
Response: 887301
Conc: 15.14 ng/ml



#3 AR-1016-1


R.T.: 5.303 min
Delta R.T.: 0.003 min
Response: 370293
Conc: 281.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#3 AR-1016-1

R.T.: 4.488 min
Delta R.T.: -0.004 min
Response: 367559
Conc: 211.21 ng/ml




#4 AR-1016-2

R.T.: 5.325 min
Delta R.T.: -0.013 min
Response: 473905
Conc: 241.34 ng/ml



#4 AR-1016-2


R.T.: 4.506 min
Delta R.T.: -0.004 min
Response: 683989
Conc: 248.03 ng/ml



#5 AR-1016-3


R.T.: 5.386 min
Delta R.T.: -0.012 min
Response: 303619
Conc: 256.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#5 AR-1016-3

R.T.: 4.678 min
Delta R.T.: -0.004 min
Response: 315451
Conc: 226.59 ng/ml




#6 AR-1016-4

R.T.: 5.483 min
Delta R.T.: -0.012 min
Response: 582095
Conc: 588.57 ng/ml



#6 AR-1016-4


R.T.: 4.720 min
Delta R.T.: -0.005 min
Response: 230789
Conc: 199.76 ng/ml



#7 AR-1016-5


R.T.: 5.772 min
Delta R.T.: -0.013 min
Response: 376750
Conc: 372.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#7 AR-1016-5

R.T.: 4.929 min
Delta R.T.: -0.004 min
Response: 306683
Conc: 205.62 ng/ml




#8 AR-1221-1

R.T.: 4.359 min
Delta R.T.: -0.003 min
Response: 23960
Conc: 67.28 ng/ml



#8 AR-1221-1


R.T.: 3.641 min
Delta R.T.: 0.003 min
Response: 25442
Conc: 50.31 ng/ml



#9 AR-1221-2


R.T.: 4.448 min
Delta R.T.: 0.002 min
Response: 38195
Conc: 145.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#9 AR-1221-2

R.T.: 3.724 min
Delta R.T.: 0.002 min
Response: 104531
Conc: 280.74 ng/ml




#10 AR-1221-3

R.T.: 4.522 min
Delta R.T.: 0.001 min
Response: 159665
Conc: 182.62 ng/ml



#10 AR-1221-3


R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 265298
Conc: 221.85 ng/ml



#11 AR-1232-1


R.T.: 4.522 min
Delta R.T.: 0.001 min
Response: 159665
Conc: 235.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#11 AR-1232-1

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 265298
Conc: 285.58 ng/ml




#12 AR-1232-2

R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 195616
Conc: 469.34 ng/ml



#12 AR-1232-2


R.T.: 4.506 min
Delta R.T.: 0.000 min
Response: 683989
Conc: 617.41 ng/ml



#13 AR-1232-3


R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 473905
Conc: 581.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#13 AR-1232-3

R.T.: 4.678 min
Delta R.T.: 0.000 min
Response: 315451
Conc: 565.83 ng/ml




#14 AR-1232-4

R.T.: 5.483 min
Delta R.T.: 0.000 min
Response: 582095
Conc: 1422.09 ng/ml



#14 AR-1232-4


R.T.: 4.761 min
Delta R.T.: 0.000 min
Response: 268557
Conc: 556.45 ng/ml



#15 AR-1232-5


R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 547788
Conc: 2000.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#15 AR-1232-5

R.T.: 4.929 min
Delta R.T.: 0.000 min
Response: 306683
Conc: 594.62 ng/ml




#16 AR-1242-1

R.T.: 5.303 min
Delta R.T.: 0.000 min
Response: 370293
Conc: 365.61 ng/ml



#16 AR-1242-1


R.T.: 4.488 min
Delta R.T.: 0.000 min
Response: 367559
Conc: 277.10 ng/ml



#17 AR-1242-2


R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 473905
Conc: 319.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#17 AR-1242-2

R.T.: 4.506 min
Delta R.T.: 0.000 min
Response: 683989
Conc: 322.57 ng/ml




#18 AR-1242-3

R.T.: 5.386 min
Delta R.T.: 0.000 min
Response: 303619
Conc: 333.18 ng/ml



#18 AR-1242-3


R.T.: 4.678 min
Delta R.T.: 0.000 min
Response: 315451
Conc: 296.78 ng/ml



#19 AR-1242-4


R.T.: 5.483 min
Delta R.T.: 0.001 min
Response: 582095
Conc: 762.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#19 AR-1242-4

R.T.: 4.761 min
Delta R.T.: 0.000 min
Response: 268557
Conc: 261.26 ng/ml




#20 AR-1242-5

R.T.: 6.210 min
Delta R.T.: 0.001 min
Response: 32125
Conc: 34.64 ng/ml



#20 AR-1242-5


R.T.: 5.275 min
Delta R.T.: 0.000 min
Response: 249510
Conc: 180.44 ng/ml



#21 AR-1248-1


R.T.: 5.303 min
Delta R.T.: 0.000 min
Response: 370293
Conc: 485.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#21 AR-1248-1

R.T.: 4.488 min
Delta R.T.: 0.000 min
Response: 367559
Conc: 353.81 ng/ml




#22 AR-1248-2

R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 547788
Conc: 505.21 ng/ml



#22 AR-1248-2


R.T.: 4.720 min
Delta R.T.: 0.000 min
Response: 230789
Conc: 155.34 ng/ml



#23 AR-1248-3


R.T.: 5.772 min
Delta R.T.: 0.000 min
Response: 376750
Conc: 308.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#23 AR-1248-3

R.T.: 4.761 min
Delta R.T.: 0.001 min
Response: 268557
Conc: 179.61 ng/ml




#24 AR-1248-4

R.T.: 6.171 min
Delta R.T.: 0.000 min
Response: 38788
Conc: 24.57 ng/ml



#24 AR-1248-4


R.T.: 4.929 min
Delta R.T.: 0.001 min
Response: 306683
Conc: 169.88 ng/ml



#25 AR-1248-5


R.T.: 6.210 min
Delta R.T.: 0.000 min
Response: 32125
Conc: 21.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#25 AR-1248-5

R.T.: 5.315 min
Delta R.T.: 0.000 min
Response: 35916
Conc: 18.37 ng/ml




#26 AR-1254-1

R.T.: 6.144 min
Delta R.T.: 0.000 min
Response: 180785
Conc: 116.36 ng/ml



#26 AR-1254-1


R.T.: 5.275 min
Delta R.T.: 0.000 min
Response: 249510
Conc: 78.39 ng/ml



#27 AR-1254-2


R.T.: 6.360 min
Delta R.T.: 0.000 min
Response: 187476
Conc: 75.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#27 AR-1254-2

R.T.: 5.421 min
Delta R.T.: 0.000 min
Response: 257129
Conc: 87.61 ng/ml




#28 AR-1254-3

R.T.: 6.723 min
Delta R.T.: 0.000 min
Response: 101308
Conc: 37.91 ng/ml



#28 AR-1254-3


R.T.: 5.833 min
Delta R.T.: 0.015 min
Response: 478063
Conc: 104.10 ng/ml



#29 AR-1254-4


R.T.: 7.005 min
Delta R.T.: 0.000 min
Response: 63400
Conc: 27.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#29 AR-1254-4

R.T.: 6.046 min
Delta R.T.: 0.000 min
Response: 53208
Conc: 16.89 ng/ml




#30 AR-1254-5

R.T.: 7.420 min
Delta R.T.: -0.001 min
Response: 667591
Conc: 292.58 ng/ml



#30 AR-1254-5


R.T.: 6.458 min
Delta R.T.: 0.000 min
Response: 994460
Conc: 234.79 ng/ml



#31 AR-1260-1


R.T.: 6.884 min
Delta R.T.: -0.012 min
Response: 443340
Conc: 221.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#31 AR-1260-1

R.T.: 5.947 min
Delta R.T.: -0.006 min
Response: 671464
Conc: 208.73 ng/ml




#32 AR-1260-2

R.T.: 7.140 min
Delta R.T.: -0.011 min
Response: 620089
Conc: 225.06 ng/ml



#32 AR-1260-2


R.T.: 6.135 min
Delta R.T.: -0.005 min
Response: 969145
Conc: 215.09 ng/ml



#33 AR-1260-3


R.T.: 7.495 min
Delta R.T.: -0.011 min
Response: 417524
Conc: 182.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#33 AR-1260-3

R.T.: 6.285 min
Delta R.T.: -0.005 min
Response: 763883
Conc: 202.65 ng/ml




#34 AR-1260-4

R.T.: 7.719 min
Delta R.T.: -0.011 min
Response: 438821
Conc: 192.51 ng/ml



#34 AR-1260-4


R.T.: 6.751 min
Delta R.T.: -0.005 min
Response: 601552
Conc: 192.22 ng/ml



#35 AR-1260-5


R.T.: 8.024 min
Delta R.T.: -0.010 min
Response: 861116
Conc: 173.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#35 AR-1260-5

R.T.: 6.994 min
Delta R.T.: -0.005 min
Response: 1593046
Conc: 192.37 ng/ml




#36 AR-1262-1

R.T.: 7.495 min
Delta R.T.: 0.000 min
Response: 417524
Conc: 126.62 ng/ml



#36 AR-1262-1


R.T.: 6.548 min
Delta R.T.: 0.000 min
Response: 408578
Conc: 193.94 ng/ml



#37 AR-1262-2


R.T.: 8.024 min
Delta R.T.: 0.000 min
Response: 861116
Conc: 157.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#37 AR-1262-2

R.T.: 6.994 min
Delta R.T.: 0.000 min
Response: 1593046
Conc: 168.34 ng/ml




#38 AR-1262-3

R.T.: 8.323 min
Delta R.T.: 0.016 min
Response: 547410
Conc: 152.12 ng/ml



#38 AR-1262-3


R.T.: 7.272 min
Delta R.T.: 0.000 min
Response: 819791
Conc: 213.01 ng/ml



#39 AR-1262-4


R.T.: 8.370 min
Delta R.T.: -0.019 min
Response: 281108
Conc: 103.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#39 AR-1262-4

R.T.: 7.336 min
Delta R.T.: 0.000 min
Response: 2588879
Conc: 393.92 ng/ml




#40 AR-1262-5

R.T.: 8.984 min
Delta R.T.: 0.000 min
Response: 166600
Conc: 87.67 ng/ml



#40 AR-1262-5


R.T.: 7.829 min
Delta R.T.: 0.000 min
Response: 262390
Conc: 93.03 ng/ml



#41 AR-1268-1


R.T.: 8.323 min
Delta R.T.: 0.017 min
Response: 547410
Conc: 79.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#41 AR-1268-1

R.T.: 7.272 min
Delta R.T.: 0.000 min
Response: 819791
Conc: 68.27 ng/ml




#42 AR-1268-2

R.T.: 8.370 min
Delta R.T.: -0.022 min
Response: 281108
Conc: 46.67 ng/ml



#42 AR-1268-2


R.T.: 7.336 min
Delta R.T.: -0.002 min
Response: 2588879
Conc: 250.67 ng/ml



#43 AR-1268-3


R.T.: 8.596 min
Delta R.T.: 0.000 min
Response: 111628
Conc: 21.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS




#43 AR-1268-3

R.T.: 7.537 min
Delta R.T.: 0.000 min
Response: 68930
Conc: 7.91 ng/ml




#44 AR-1268-4

R.T.: 8.984 min
Delta R.T.: 0.000 min
Response: 166600
Conc: 78.78 ng/ml



#44 AR-1268-4


R.T.: 7.829 min
Delta R.T.: 0.000 min
Response: 262390
Conc: 79.44 ng/ml



#45 AR-1268-5

R.T.: 9.352 min
Delta R.T.: 0.000 min
Response: 60066
Conc: 3.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB126560BS



#45 AR-1268-5

R.T.: 8.098 min
Delta R.T.: 0.000 min
Response: 92737
Conc: 3.74 ng/ml