

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020420\
 Data File : P0066257.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Feb 2020 21:22
 Operator : DD\AJ
 Sample : L1376-02 5X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 02-03-B

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 04:45:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 31 01:14:46 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.154	3.430	153402	228817	5.266	5.594
2) SA Decachlor...	9.648	8.326	145627	224745	3.695	3.835

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020420\
Data File : PO066257.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Feb 2020 21:22
Operator : DD\AJ
Sample : L1376-02 5X
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
02-03-B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 04:45:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO013020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 31 01:14:46 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.154 min
Delta R.T.: -0.011 min
Response: 153402
Conc: 5.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
02-03-B




#1 Tetrachloro-m-xylene

R.T.: 3.430 min
Delta R.T.: -0.003 min
Response: 228817
Conc: 5.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.648 min
Delta R.T.: -0.013 min
Response: 145627
Conc: 3.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.326 min
Delta R.T.: -0.006 min
Response: 224745
Conc: 3.83 ng/ml