

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020421\
 Data File : P0075510.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Feb 2021 13:58
 Operator : AJ/MA
 Sample : M1291-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 OU421-007-0002

Manual Integrations
 APPROVED

Ankita
 2/8/2021 8:40:42 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 01:05:49 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 09:09:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.922	3.994	2175989	1112188	21.501m	23.136
2) SA Decachlor...	10.936	9.242	4420484	1870039	45.817	38.525
Target Compounds						
36) L8 AR-1262-1	8.572	7.340	10538025	4279925	1207.591	1962.171 #
37) L8 AR-1262-2	9.130	7.888	11742436	4858526	776.515	633.528
38) L8 AR-1262-3	9.437	8.171	25966962	9840767	2485.628	2973.648m
39) L8 AR-1262-4	9.534	8.236	23942016	10299087	3239.288	1793.697m#
40) L8 AR-1262-5	10.196	8.729	14507400	5939961	2797.466	2159.317

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020421\
Data File : PO075510.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Feb 2021 13:58
Operator : AJ/MA
Sample : M1291-01
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
OU421-007-0002


Manual Integrations
APPROVED

Ankita
2/8/2021 8:40:42 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 01:05:49 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 09:09:46 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.922 min
Delta R.T.: -0.002 min
Response: 2175989
Conc: 21.50 ng/ml m

Instrument :
ECD_O
ClientSampleId :
OU421-007-0002


Manual Integrations
APPROVED

Ankita
2/8/2021 8:40:42 AM




#1 Tetrachloro-m-xylene

R.T.: 3.994 min
Delta R.T.: -0.003 min
Response: 1112188
Conc: 23.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.936 min
Delta R.T.: -0.002 min
Response: 4420484
Conc: 45.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.242 min
Delta R.T.: -0.004 min
Response: 1870039
Conc: 38.52 ng/ml




#36 AR-1262-1

R.T.: 8.572 min
Delta R.T.: 0.000 min
Response: 10538025
Conc: 1207.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU421-007-0002


Manual Integrations
APPROVED

Ankita
2/8/2021 8:40:42 AM




#36 AR-1262-1

R.T.: 7.340 min
Delta R.T.: -0.001 min
Response: 4279925
Conc: 1962.17 ng/ml




#37 AR-1262-2

R.T.: 9.130 min
Delta R.T.: 0.000 min
Response: 11742436
Conc: 776.51 ng/ml



#37 AR-1262-2

R.T.: 7.888 min
Delta R.T.: -0.002 min
Response: 4858526
Conc: 633.53 ng/ml




#38 AR-1262-3

R.T.: 9.437 min
Delta R.T.: 0.000 min
Response: 25966962
Conc: 2485.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU421-007-0002


Manual Integrations
APPROVED

Ankita
2/8/2021 8:40:42 AM




#38 AR-1262-3

R.T.: 8.171 min
Delta R.T.: -0.003 min
Response: 9840767
Conc: 2973.65 ng/ml m




#39 AR-1262-4

R.T.: 9.534 min
Delta R.T.: 0.003 min
Response: 23942016
Conc: 3239.29 ng/ml



#39 AR-1262-4

R.T.: 8.236 min
Delta R.T.: -0.001 min
Response: 10299087
Conc: 1793.70 ng/ml m




#40 AR-1262-5

R.T.: 10.196 min
Delta R.T.: 0.000 min
Response: 14507400
Conc: 2797.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
OU421-007-0002

Manual Integrations
APPROVED

Ankita
2/8/2021 8:40:42 AM



#40 AR-1262-5

R.T.: 8.729 min
Delta R.T.: -0.002 min
Response: 5939961
Conc: 2159.32 ng/ml