

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0020521\
 Data File : P0075595.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Feb 2021 00:10
 Operator : AJ/MA
 Sample : M1292-08
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 OU421-003-1015

Manual Integrations
 APPROVED

Ankita
 2/8/2021 4:51:42 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 00:39:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 09:09:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.921	3.993	2379231	1178269	23.510	24.511
2) SA Decachlor...	10.933	9.240	3370046	1640985	34.929	33.806
Target Compounds						
36) L8 AR-1262-1	8.568	7.337	170226	121753	19.507m	55.819 #
37) L8 AR-1262-2	9.125	7.885	184791	141732	12.220m	18.481 #
38) L8 AR-1262-3	9.435	8.168	329220	173536	31.514	52.439 #
39) L8 AR-1262-4	9.530	8.234	328314	184171	44.420	32.075 #
40) L8 AR-1262-5	10.193	8.726	230363	105817	44.421	38.467

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020521\
Data File : PO075595.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Feb 2021 00:10
Operator : AJ/MA
Sample : M1292-08
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
OU421-003-1015

Manual Integrations
APPROVED

Ankita
2/8/2021 4:51:42 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 00:39:22 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 09:09:46 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

