

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020521\
 Data File : P0075575.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Feb 2021 15:35
 Operator : AJ/MA
 Sample : AR1262CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262CCC500

Manual Integrations
 APPROVED

Ankita
 2/8/2021 4:50:43 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 17:27:06 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 09:09:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.924	3.995	5510909	2645124	54.455	55.025
2) SA Decachlor...	10.939	9.243	5244055	2445873	54.353	50.387
Target Compounds						
36) L8 AR-1262-1	8.573	7.339	4230744	1038270	484.816	476.005
37) L8 AR-1262-2	9.131	7.889	7242716	3519818	478.953	458.967
38) L8 AR-1262-3	9.440	8.172	4817255	1490170	461.121	450.294
39) L8 AR-1262-4	9.533	8.236	3564339	2607953	482.245m	454.203
40) L8 AR-1262-5	10.197	8.730	2580192	1227018	497.539	446.050

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020521\
Data File : PO075575.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Feb 2021 15:35
Operator : AJ/MA
Sample : AR1262CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1262CCC500


Manual Integrations
APPROVED

Ankita
2/8/2021 4:50:43 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 17:27:06 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 09:09:46 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.924 min
Delta R.T.: 0.000 min
Response: 5510909
Conc: 54.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262CCC500


Manual Integrations
APPROVED

Ankita
2/8/2021 4:50:43 PM




#1 Tetrachloro-m-xylene

R.T.: 3.995 min
Delta R.T.: -0.001 min
Response: 2645124
Conc: 55.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.939 min
Delta R.T.: 0.002 min
Response: 5244055
Conc: 54.35 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.243 min
Delta R.T.: -0.003 min
Response: 2445873
Conc: 50.39 ng/ml




#36 AR-1262-1

R.T.: 8.573 min
Delta R.T.: 0.000 min
Response: 4230744
Conc: 484.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262CCC500


Manual Integrations
APPROVED

Ankita
2/8/2021 4:50:43 PM




#36 AR-1262-1

R.T.: 7.339 min
Delta R.T.: -0.002 min
Response: 1038270
Conc: 476.00 ng/ml




#37 AR-1262-2

R.T.: 9.131 min
Delta R.T.: 0.002 min
Response: 7242716
Conc: 478.95 ng/ml



#37 AR-1262-2

R.T.: 7.889 min
Delta R.T.: -0.001 min
Response: 3519818
Conc: 458.97 ng/ml




#38 AR-1262-3

R.T.: 9.440 min
Delta R.T.: 0.002 min
Response: 4817255
Conc: 461.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262CCC500


Manual Integrations
APPROVED

Ankita
2/8/2021 4:50:43 PM




#38 AR-1262-3

R.T.: 8.172 min
Delta R.T.: -0.002 min
Response: 1490170
Conc: 450.29 ng/ml




#39 AR-1262-4

R.T.: 9.533 min
Delta R.T.: 0.002 min
Response: 3564339
Conc: 482.25 ng/ml m



#39 AR-1262-4

R.T.: 8.236 min
Delta R.T.: -0.002 min
Response: 2607953
Conc: 454.20 ng/ml




#40 AR-1262-5

R.T.: 10.197 min
Delta R.T.: 0.002 min
Response: 2580192
Conc: 497.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262CCC500

Manual Integrations
APPROVED

Ankita
2/8/2021 4:50:43 PM



#40 AR-1262-5

R.T.: 8.730 min
Delta R.T.: -0.001 min
Response: 1227018
Conc: 446.05 ng/ml