

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020521\
 Data File : P0075613.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Feb 2021 5:52
 Operator : AJ/MA
 Sample : PB134474BL
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB134474BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 06:11:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 09:09:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.921	3.994	1175942	976754	11.620	20.319 #
2)	SA Decachlor...	0.000	9.244	0	221540	N.D.	4.564 #
Target Compounds							
26)	L6 AR-1254-1	7.159	6.102	124.6E6	84520024	28900.517	27289.364
27)	L6 AR-1254-2	7.389	6.261	260.2E6	82836071	38598.813	30847.528
28)	L6 AR-1254-3	7.772	6.677	280.6E6	144.0E6	40683.730	33765.678
29)	L6 AR-1254-4	8.069	6.915	206.6E6	101.6E6	34173.458	34863.546
30)	L6 AR-1254-5	8.499	7.340	306.3E6	165.0E6	54802.158	42502.116



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020521\
Data File : P0075613.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Feb 2021 5:52
Operator : AJ/MA
Sample : PB134474BL
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB134474BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 06:11:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 09:09:46 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.921 min
Delta R.T.: -0.003 min
Response: 1175942
Conc: 11.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB134474BL




#1 Tetrachloro-m-xylene

R.T.: 3.994 min
Delta R.T.: -0.002 min
Response: 976754
Conc: 20.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.938 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 9.244 min
Delta R.T.: -0.002 min
Response: 221540
Conc: 4.56 ng/ml



#26 AR-1254-1


R.T.: 7.159 min
Delta R.T.: -0.001 min
Response: 124557632
Conc: 28900.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB134474BL




#26 AR-1254-1

R.T.: 6.102 min
Delta R.T.: -0.002 min
Response: 84520024
Conc: 27289.36 ng/ml




#27 AR-1254-2

R.T.: 7.389 min
Delta R.T.: -0.002 min
Response: 260239110
Conc: 38598.81 ng/ml



#27 AR-1254-2


R.T.: 6.261 min
Delta R.T.: -0.003 min
Response: 82836071
Conc: 30847.53 ng/ml



#28 AR-1254-3


R.T.: 7.772 min
Delta R.T.: 0.000 min
Response: 280559546
Conc: 40683.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB134474BL




#28 AR-1254-3

R.T.: 6.677 min
Delta R.T.: -0.002 min
Response: 143976825
Conc: 33765.68 ng/ml




#29 AR-1254-4

R.T.: 8.069 min
Delta R.T.: 0.000 min
Response: 206596257
Conc: 34173.46 ng/ml



#29 AR-1254-4


R.T.: 6.915 min
Delta R.T.: -0.003 min
Response: 101616567
Conc: 34863.55 ng/ml



#30 AR-1254-5

R.T.: 8.499 min
Delta R.T.: 0.000 min
Response: 306337442
Conc: 54802.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB134474BL



#30 AR-1254-5

R.T.: 7.340 min
Delta R.T.: -0.003 min
Response: 164956198
Conc: 42502.12 ng/ml