

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020623\
 Data File : P0092375.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2023 10:11
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 16:26:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 04:40:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.449	3.644	137.2E6	64749229	47.407	46.113
2) SA Decachlor...	10.378	8.686	100.7E6	51946620	48.774	46.015
Target Compounds						
3) L1 AR-1016-1	5.636	4.732	45082505	21036047	492.934	457.697
4) L1 AR-1016-2	5.659	4.750	65676667	30545159	490.455	465.785
5) L1 AR-1016-3	5.722	4.927	41594244	16208877	483.348	458.462
6) L1 AR-1016-4	5.822	4.969	32256921	14195800	483.860	456.354
7) L1 AR-1016-5	6.123	5.183	34705652	18204249	483.027	457.784
31) L7 AR-1260-1	7.276	6.222	65109718	33473553	494.170	443.126
32) L7 AR-1260-2	7.540	6.411	72355331	38516044	487.942	439.612
33) L7 AR-1260-3	7.908	6.565	55469075	35348752	489.735	407.764
34) L7 AR-1260-4	8.138	7.040	60940545	29936176	491.595	447.025
35) L7 AR-1260-5	8.474	7.283	111.2E6	64955785	500.765	456.470



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020623\
Data File : PO092375.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2023 10:11
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 16:26:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 04:40:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: 0.000 min
Response: 137239399
Conc: 47.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: -0.002 min
Response: 64749229
Conc: 46.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.378 min
Delta R.T.: 0.007 min
Response: 100667284
Conc: 48.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.686 min
Delta R.T.: -0.008 min
Response: 51946620
Conc: 46.02 ng/ml



#3 AR-1016-1


R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 45082505
Conc: 492.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.732 min
Delta R.T.: -0.003 min
Response: 21036047
Conc: 457.70 ng/ml




#4 AR-1016-2

R.T.: 5.659 min
Delta R.T.: 0.001 min
Response: 65676667
Conc: 490.45 ng/ml



#4 AR-1016-2


R.T.: 4.750 min
Delta R.T.: -0.003 min
Response: 30545159
Conc: 465.79 ng/ml



#5 AR-1016-3


R.T.: 5.722 min
Delta R.T.: 0.001 min
Response: 41594244
Conc: 483.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.927 min
Delta R.T.: -0.004 min
Response: 16208877
Conc: 458.46 ng/ml




#6 AR-1016-4

R.T.: 5.822 min
Delta R.T.: 0.000 min
Response: 32256921
Conc: 483.86 ng/ml



#6 AR-1016-4


R.T.: 4.969 min
Delta R.T.: -0.003 min
Response: 14195800
Conc: 456.35 ng/ml



#7 AR-1016-5


R.T.: 6.123 min
Delta R.T.: 0.001 min
Response: 34705652
Conc: 483.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.183 min
Delta R.T.: -0.004 min
Response: 18204249
Conc: 457.78 ng/ml




#31 AR-1260-1

R.T.: 7.276 min
Delta R.T.: 0.004 min
Response: 65109718
Conc: 494.17 ng/ml



#31 AR-1260-1


R.T.: 6.222 min
Delta R.T.: -0.004 min
Response: 33473553
Conc: 443.13 ng/ml



#32 AR-1260-2


R.T.: 7.540 min
Delta R.T.: 0.003 min
Response: 72355331
Conc: 487.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.411 min
Delta R.T.: -0.004 min
Response: 38516044
Conc: 439.61 ng/ml




#33 AR-1260-3

R.T.: 7.908 min
Delta R.T.: 0.004 min
Response: 55469075
Conc: 489.73 ng/ml



#33 AR-1260-3


R.T.: 6.565 min
Delta R.T.: -0.005 min
Response: 35348752
Conc: 407.76 ng/ml



#34 AR-1260-4


R.T.: 8.138 min
Delta R.T.: 0.004 min
Response: 60940545
Conc: 491.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.040 min
Delta R.T.: -0.006 min
Response: 29936176
Conc: 447.02 ng/ml



#35 AR-1260-5

R.T.: 8.474 min
Delta R.T.: 0.005 min
Response: 111198581
Conc: 500.77 ng/ml



#35 AR-1260-5

R.T.: 7.283 min
Delta R.T.: -0.005 min
Response: 64955785
Conc: 456.47 ng/ml