

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020822\
 Data File : P0084520.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Feb 2022 21:04
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 10 06:10:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0020822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Feb 10 06:10:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.764	672001	257058	5.703	5.558
2) SA Decachlor...	10.445	8.806	434564	247229	5.114	6.211
Target Compounds						
26) L6 AR-1254-1	6.570	5.661	242314	174890	54.859	63.646
27) L6 AR-1254-2	6.791	5.809	358389	153960	54.236	64.226
28) L6 AR-1254-3	7.163	6.213	363497	233366	53.553	62.673
29) L6 AR-1254-4	7.453	6.441	253832	142389	53.869	62.584
30) L6 AR-1254-5	7.878	6.860	301915	204166	56.876	62.590



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020822\
Data File : P0084520.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Feb 2022 21:04
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 10 06:10:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0020822.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 10 06:10:12 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.507 min
Delta R.T.: 0.000 min
Response: 672001
Conc: 5.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.764 min
Delta R.T.: 0.000 min
Response: 257058
Conc: 5.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.445 min
Delta R.T.: 0.000 min
Response: 434564
Conc: 5.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 247229
Conc: 6.21 ng/ml



#26 AR-1254-1


R.T.: 6.570 min
Delta R.T.: 0.000 min
Response: 242314
Conc: 54.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#26 AR-1254-1

R.T.: 5.661 min
Delta R.T.: 0.000 min
Response: 174890
Conc: 63.65 ng/ml




#27 AR-1254-2

R.T.: 6.791 min
Delta R.T.: 0.000 min
Response: 358389
Conc: 54.24 ng/ml



#27 AR-1254-2


R.T.: 5.809 min
Delta R.T.: 0.000 min
Response: 153960
Conc: 64.23 ng/ml



#28 AR-1254-3


R.T.: 7.163 min
Delta R.T.: 0.000 min
Response: 363497
Conc: 53.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#28 AR-1254-3

R.T.: 6.213 min
Delta R.T.: 0.000 min
Response: 233366
Conc: 62.67 ng/ml




#29 AR-1254-4

R.T.: 7.453 min
Delta R.T.: 0.000 min
Response: 253832
Conc: 53.87 ng/ml



#29 AR-1254-4


R.T.: 6.441 min
Delta R.T.: 0.000 min
Response: 142389
Conc: 62.58 ng/ml



#30 AR-1254-5

R.T.: 7.878 min
Delta R.T.: 0.000 min
Response: 301915
Conc: 56.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050



#30 AR-1254-5

R.T.: 6.860 min
Delta R.T.: 0.000 min
Response: 204166
Conc: 62.59 ng/ml