

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020823\
 Data File : P0092416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Feb 2023 15:48
 Operator : YP/AJ
 Sample : 01481-11
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 JPP-46-020723

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 02/09/2023
 Supervised By :Sohil Jodhani 02/09/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 09 09:32:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 04:40:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.449	3.641	67205305	29303292	23.215	20.869
2) SA Decachlor...	10.381	8.682	54274089	25161928	26.296	22.289
Target Compounds						
21) L5 AR-1248-1	5.637	4.728	5246896	1799806	90.394m	61.780 #
22) L5 AR-1248-2	5.918	4.966	6023671	2813424	65.916	62.358
23) L5 AR-1248-3	6.124	5.009	5970849	2265179	60.881m	48.568
24) L5 AR-1248-4	6.538	5.178	6687638	2785949	68.305	51.851
25) L5 AR-1248-5	6.579	5.574	8776601	2274466	93.091	48.043m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020823\
Data File : PO092416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Feb 2023 15:48
Operator : YP/AJ
Sample : 01481-11
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
JPP-46-020723


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/09/2023
Supervised By :Sohil Jodhani 02/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 09:32:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 04:40:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.449 min
Delta R.T.: 0.000 min
Response: 67205305
Conc: 23.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
JPP-46-020723


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/09/2023
Supervised By :Sohil Jodhani 02/09/2023




#1 Tetrachloro-m-xylene

R.T.: 3.641 min
Delta R.T.: -0.005 min
Response: 29303292
Conc: 20.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.381 min
Delta R.T.: 0.010 min
Response: 54274089
Conc: 26.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.682 min
Delta R.T.: -0.011 min
Response: 25161928
Conc: 22.29 ng/ml




#21 AR-1248-1

R.T.: 5.637 min
Delta R.T.: 0.002 min
Response: 5246896
Conc: 90.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
JPP-46-020723


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/09/2023
Supervised By :Sohil Jodhani 02/09/2023




#21 AR-1248-1

R.T.: 4.728 min
Delta R.T.: -0.006 min
Response: 1799806
Conc: 61.78 ng/ml




#22 AR-1248-2

R.T.: 5.918 min
Delta R.T.: 0.003 min
Response: 6023671
Conc: 65.92 ng/ml



#22 AR-1248-2

R.T.: 4.966 min
Delta R.T.: -0.006 min
Response: 2813424
Conc: 62.36 ng/ml




#23 AR-1248-3

R.T.: 6.124 min
Delta R.T.: 0.002 min
Response: 5970849
Conc: 60.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
JPP-46-020723


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/09/2023
Supervised By :Sohil Jodhani 02/09/2023




#23 AR-1248-3

R.T.: 5.009 min
Delta R.T.: -0.005 min
Response: 2265179
Conc: 48.57 ng/ml




#24 AR-1248-4

R.T.: 6.538 min
Delta R.T.: 0.005 min
Response: 6687638
Conc: 68.30 ng/ml



#24 AR-1248-4

R.T.: 5.178 min
Delta R.T.: -0.008 min
Response: 2785949
Conc: 51.85 ng/ml




#25 AR-1248-5

R.T.: 6.579 min
Delta R.T.: 0.006 min
Response: 8776601
Conc: 93.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
JPP-46-020723

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/09/2023
Supervised By :Sohil Jodhani 02/09/2023



#25 AR-1248-5

R.T.: 5.574 min
Delta R.T.: -0.006 min
Response: 2274466
Conc: 48.04 ng/ml m