

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021022\
 Data File : P0084567.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Feb 2022 11:11
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 10 16:40:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0020822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Feb 10 07:10:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.507	3.779	5972328	2242240	52.525	48.803
2)	SA Decachlor...	10.447	8.824	4343323	1832080	54.226	48.918
Target Compounds							
26)	L6 AR-1254-1	6.570	5.676	2252581	1241459	509.976	451.792
27)	L6 AR-1254-2	6.792	5.823	3336978	1080772	504.999	450.854
28)	L6 AR-1254-3	7.164	6.228	3447834	1646067	507.959	442.066
29)	L6 AR-1254-4	7.454	6.456	2441946	1059914	518.240	465.863
30)	L6 AR-1254-5	7.880	6.874	2968294	1498361	559.182	459.346



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021022\
Data File : P0084567.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Feb 2022 11:11
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 10 16:40:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0020822.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 10 07:10:18 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.507 min
Delta R.T.: 0.000 min
Response: 5972328
Conc: 52.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: 0.019 min
Response: 2242240
Conc: 48.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.447 min
Delta R.T.: 0.000 min
Response: 4343323
Conc: 54.23 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.824 min
Delta R.T.: 0.023 min
Response: 1832080
Conc: 48.92 ng/ml



#26 AR-1254-1


R.T.: 6.570 min
Delta R.T.: 0.000 min
Response: 2252581
Conc: 509.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.676 min
Delta R.T.: 0.016 min
Response: 1241459
Conc: 451.79 ng/ml




#27 AR-1254-2

R.T.: 6.792 min
Delta R.T.: 0.000 min
Response: 3336978
Conc: 505.00 ng/ml



#27 AR-1254-2


R.T.: 5.823 min
Delta R.T.: 0.015 min
Response: 1080772
Conc: 450.85 ng/ml



#28 AR-1254-3


R.T.: 7.164 min
Delta R.T.: 0.000 min
Response: 3447834
Conc: 507.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.228 min
Delta R.T.: 0.015 min
Response: 1646067
Conc: 442.07 ng/ml




#29 AR-1254-4

R.T.: 7.454 min
Delta R.T.: 0.000 min
Response: 2441946
Conc: 518.24 ng/ml



#29 AR-1254-4


R.T.: 6.456 min
Delta R.T.: 0.015 min
Response: 1059914
Conc: 465.86 ng/ml



#30 AR-1254-5

R.T.: 7.880 min
Delta R.T.: 0.001 min
Response: 2968294
Conc: 559.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.874 min
Delta R.T.: 0.015 min
Response: 1498361
Conc: 459.35 ng/ml