

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021024\
 Data File : P0101666.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Feb 2024 05:10
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 04:35:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 12 04:34:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.290	3.400	91289703	65958753	23.709	25.455
2) SA Decachlor...	9.844	8.284	56147777	47935534	26.081	25.612
Target Compounds						
26) L6 AR-1254-1	6.277	5.241	29769195	30677298	266.766	260.582
27) L6 AR-1254-2	6.493	5.387	43772791	27857994	266.578	264.338
28) L6 AR-1254-3	6.855	5.784	43024119	42083039	274.743	258.385
29) L6 AR-1254-4	7.137	6.010	25926126	21805349	262.874	257.260
30) L6 AR-1254-5	7.553	6.422	30921776	34451309	265.996	253.032


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO021024\
Data File : PO101666.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Feb 2024 05:10
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 04:35:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO021024.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 12 04:34:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.290 min
Delta R.T.: 0.000 min
Response: 91289703
Conc: 23.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 65958753
Conc: 25.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.844 min
Delta R.T.: 0.000 min
Response: 56147777
Conc: 26.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.284 min
Delta R.T.: 0.000 min
Response: 47935534
Conc: 25.61 ng/ml



#26 AR-1254-1


R.T.: 6.277 min
Delta R.T.: 0.000 min
Response: 29769195
Conc: 266.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254\ICC250




#26 AR-1254-1

R.T.: 5.241 min
Delta R.T.: 0.000 min
Response: 30677298
Conc: 260.58 ng/ml




#27 AR-1254-2

R.T.: 6.493 min
Delta R.T.: 0.000 min
Response: 43772791
Conc: 266.58 ng/ml



#27 AR-1254-2


R.T.: 5.387 min
Delta R.T.: 0.000 min
Response: 27857994
Conc: 264.34 ng/ml



#28 AR-1254-3


R.T.: 6.855 min
Delta R.T.: 0.000 min
Response: 43024119
Conc: 274.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 42083039
Conc: 258.38 ng/ml




#29 AR-1254-4

R.T.: 7.137 min
Delta R.T.: 0.000 min
Response: 25926126
Conc: 262.87 ng/ml



#29 AR-1254-4


R.T.: 6.010 min
Delta R.T.: 0.000 min
Response: 21805349
Conc: 257.26 ng/ml



#30 AR-1254-5

R.T.: 7.553 min
Delta R.T.: 0.000 min
Response: 30921776
Conc: 266.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.422 min
Delta R.T.: 0.000 min
Response: 34451309
Conc: 253.03 ng/ml