

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021024\
 Data File : P0101679.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Feb 2024 09:48
 Operator : YP/AJ
 Sample : P0021024ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO021024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 10:01:40 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 12 05:22:08 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.290	3.398	166.8E6	118.6E6	54.103	44.609
2) SA Decachlor...	9.848	8.279	110.7E6	86426471	52.772	45.491
Target Compounds						
3) L1 AR-1016-1	5.437	4.453	51183468	32933615	519.009	451.392
4) L1 AR-1016-2	5.459	4.471	74656496	47627378	513.525	451.441
5) L1 AR-1016-3	5.520	4.643	48976486	26242533	514.611	457.435
6) L1 AR-1016-4	5.617	4.684	38785054	23081239	508.397	458.490
7) L1 AR-1016-5	5.907	4.893	38633999	28951806	486.698	457.775
31) L7 AR-1260-1	7.019	5.907	66001871	53855630	502.843	436.211
32) L7 AR-1260-2	7.273	6.094	72306236	63389986	507.313	439.110
33) L7 AR-1260-3	7.628	6.245	53247611	58202169	505.677	445.310
34) L7 AR-1260-4	7.852	6.710	57525656	47816045	515.016	433.557
35) L7 AR-1260-5	8.157	6.951	108.5E6	103.6E6	494.549	438.233



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO021024\
Data File : PO101679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Feb 2024 09:48
Operator : YP/AJ
Sample : PO021024ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO021024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 10:01:40 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO021024.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 12 05:22:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.290 min
Delta R.T.: -0.001 min
Response: 166827604
Conc: 54.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO021024




#1 Tetrachloro-m-xylene

R.T.: 3.398 min
Delta R.T.: -0.002 min
Response: 118578072
Conc: 44.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.848 min
Delta R.T.: 0.002 min
Response: 110660093
Conc: 52.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.279 min
Delta R.T.: -0.006 min
Response: 86426471
Conc: 45.49 ng/ml



#3 AR-1016-1


R.T.: 5.437 min
Delta R.T.: 0.000 min
Response: 51183468
Conc: 519.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO021024




#3 AR-1016-1

R.T.: 4.453 min
Delta R.T.: -0.002 min
Response: 32933615
Conc: 451.39 ng/ml




#4 AR-1016-2

R.T.: 5.459 min
Delta R.T.: 0.000 min
Response: 74656496
Conc: 513.52 ng/ml



#4 AR-1016-2


R.T.: 4.471 min
Delta R.T.: -0.003 min
Response: 47627378
Conc: 451.44 ng/ml



#5 AR-1016-3


R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 48976486
Conc: 514.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO021024




#5 AR-1016-3

R.T.: 4.643 min
Delta R.T.: -0.002 min
Response: 26242533
Conc: 457.44 ng/ml




#6 AR-1016-4

R.T.: 5.617 min
Delta R.T.: 0.000 min
Response: 38785054
Conc: 508.40 ng/ml



#6 AR-1016-4


R.T.: 4.684 min
Delta R.T.: -0.002 min
Response: 23081239
Conc: 458.49 ng/ml



#7 AR-1016-5


R.T.: 5.907 min
Delta R.T.: 0.000 min
Response: 38633999
Conc: 486.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO021024




#7 AR-1016-5

R.T.: 4.893 min
Delta R.T.: -0.002 min
Response: 28951806
Conc: 457.77 ng/ml




#31 AR-1260-1

R.T.: 7.019 min
Delta R.T.: 0.000 min
Response: 66001871
Conc: 502.84 ng/ml



#31 AR-1260-1


R.T.: 5.907 min
Delta R.T.: -0.003 min
Response: 53855630
Conc: 436.21 ng/ml



#32 AR-1260-2


R.T.: 7.273 min
Delta R.T.: 0.000 min
Response: 72306236
Conc: 507.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO021024




#32 AR-1260-2

R.T.: 6.094 min
Delta R.T.: -0.002 min
Response: 63389986
Conc: 439.11 ng/ml




#33 AR-1260-3

R.T.: 7.628 min
Delta R.T.: 0.000 min
Response: 53247611
Conc: 505.68 ng/ml



#33 AR-1260-3


R.T.: 6.245 min
Delta R.T.: -0.003 min
Response: 58202169
Conc: 445.31 ng/ml



#34 AR-1260-4


R.T.: 7.852 min
Delta R.T.: 0.000 min
Response: 57525656
Conc: 515.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO021024




#34 AR-1260-4

R.T.: 6.710 min
Delta R.T.: -0.003 min
Response: 47816045
Conc: 433.56 ng/ml



#35 AR-1260-5

R.T.: 8.157 min
Delta R.T.: 0.000 min
Response: 108502932
Conc: 494.55 ng/ml



#35 AR-1260-5

R.T.: 6.951 min
Delta R.T.: -0.003 min
Response: 103591375
Conc: 438.23 ng/ml