

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021420\
 Data File : P0066556.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Feb 2020 17:04
 Operator : DD\AJ
 Sample : L1523-03
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 COMP-3

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 22:23:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Feb 13 16:40:54 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.145	3.423	814815	961378	23.809	23.865
2) SA Decachlor...	9.627	8.310	942545	1178109	20.553	19.590

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021420\
Data File : P0066556.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Feb 2020 17:04
Operator : DD\AJ
Sample : L1523-03
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
COMP-3

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 22:23:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021320.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 13 16:40:54 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.145 min
Delta R.T.: 0.010 min
Response: 814815
Conc: 23.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-3




#1 Tetrachloro-m-xylene

R.T.: 3.423 min
Delta R.T.: 0.011 min
Response: 961378
Conc: 23.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.627 min
Delta R.T.: 0.004 min
Response: 942545
Conc: 20.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.310 min
Delta R.T.: 0.007 min
Response: 1178109
Conc: 19.59 ng/ml