

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021422\
 Data File : P0084625.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Feb 2022 13:49
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 14:31:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 14 14:10:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.597	9887351	2351518	50.000	50.000
2) SA Decachlor...	10.432	8.626	6477076	1866255	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.697	4.688	1861623	486626	500.000	500.000
22) L5 AR-1248-2	5.976	4.926	2613601	673925	500.000	500.000
23) L5 AR-1248-3	6.183	4.968	2872721	707710	500.000	500.000
24) L5 AR-1248-4	6.592	5.139	3199585	828701	500.000	500.000
25) L5 AR-1248-5	6.632	5.534	3058616	812728	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021422\
Data File : P0084625.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Feb 2022 13:49
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 14:31:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 14 14:10:42 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.507 min
Delta R.T.: 0.000 min
Response: 9887351
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.597 min
Delta R.T.: 0.000 min
Response: 2351518
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.432 min
Delta R.T.: 0.000 min
Response: 6477076
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.626 min
Delta R.T.: 0.000 min
Response: 1866255
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 5.697 min
Delta R.T.: 0.000 min
Response: 1861623
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 486626
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.976 min
Delta R.T.: 0.000 min
Response: 2613601
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 4.926 min
Delta R.T.: 0.000 min
Response: 673925
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 6.183 min
Delta R.T.: 0.000 min
Response: 2872721
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 4.968 min
Delta R.T.: 0.000 min
Response: 707710
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.592 min
Delta R.T.: 0.000 min
Response: 3199585
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.139 min
Delta R.T.: 0.000 min
Response: 828701
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.632 min
Delta R.T.: 0.000 min
Response: 3058616
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 812728
Conc: 500.00 ng/ml