

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021422\
 Data File : P0084638.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Feb 2022 17:31
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO021422AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 17:54:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 14 17:50:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.506	3.597	9910501	2385809	50.404	52.902
2) SA Decachlor...	10.429	8.625	6597720	1805697	53.170	49.533
Target Compounds						
26) L6 AR-1254-1	6.565	5.493	3248593	1282462	502.209	501.046
27) L6 AR-1254-2	6.787	5.641	4954635	1119355	501.502	500.105
28) L6 AR-1254-3	7.159	6.046	5228498	1817015	504.711	499.182
29) L6 AR-1254-4	7.448	6.276	3745116	1125204	511.536	500.779
30) L6 AR-1254-5	7.872	6.696	3950838	1599493	490.816	489.308


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021422\
Data File : P0084638.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Feb 2022 17:31
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO021422AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 17:54:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 14 17:50:29 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.506 min
Delta R.T.: 0.000 min
Response: 9910501
Conc: 50.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO021422AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.597 min
Delta R.T.: 0.000 min
Response: 2385809
Conc: 52.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.429 min
Delta R.T.: -0.002 min
Response: 6597720
Conc: 53.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.625 min
Delta R.T.: 0.000 min
Response: 1805697
Conc: 49.53 ng/ml



#26 AR-1254-1


R.T.: 6.565 min
Delta R.T.: 0.000 min
Response: 3248593
Conc: 502.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO021422AR1254




#26 AR-1254-1

R.T.: 5.493 min
Delta R.T.: 0.000 min
Response: 1282462
Conc: 501.05 ng/ml




#27 AR-1254-2

R.T.: 6.787 min
Delta R.T.: 0.000 min
Response: 4954635
Conc: 501.50 ng/ml



#27 AR-1254-2


R.T.: 5.641 min
Delta R.T.: -0.001 min
Response: 1119355
Conc: 500.11 ng/ml



#28 AR-1254-3


R.T.: 7.159 min
Delta R.T.: 0.000 min
Response: 5228498
Conc: 504.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO021422AR1254




#28 AR-1254-3

R.T.: 6.046 min
Delta R.T.: 0.000 min
Response: 1817015
Conc: 499.18 ng/ml




#29 AR-1254-4

R.T.: 7.448 min
Delta R.T.: 0.000 min
Response: 3745116
Conc: 511.54 ng/ml



#29 AR-1254-4


R.T.: 6.276 min
Delta R.T.: -0.001 min
Response: 1125204
Conc: 500.78 ng/ml



#30 AR-1254-5

R.T.: 7.872 min
Delta R.T.: -0.002 min
Response: 3950838
Conc: 490.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO021422AR1254



#30 AR-1254-5

R.T.: 6.696 min
Delta R.T.: -0.001 min
Response: 1599493
Conc: 489.31 ng/ml