

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021522\
 Data File : P0084655.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Feb 2022 16:12
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 01:08:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 14 17:50:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.509	3.597	9686508	2305333	49.265	51.118
2) SA Decachlor...	10.428	8.623	5965657	1701243	48.076	46.668
Target Compounds						
3) L1 AR-1016-1	5.698	4.688	3002524	835419	463.752	486.626
4) L1 AR-1016-2	5.721	4.706	4313435	1202204	463.976	496.747
5) L1 AR-1016-3	5.785	4.882	2572652	636962	457.402	484.696
6) L1 AR-1016-4	5.885	4.926	2060679	532079	444.184	480.234
7) L1 AR-1016-5	6.184	5.139	1960371	638815	437.526	461.561
31) L7 AR-1260-1	7.326	6.176	3128164	1126506	452.002	464.924
32) L7 AR-1260-2	7.586	6.366	3741965	1318523	457.774	455.197
33) L7 AR-1260-3	7.950	6.519	2607422	1236903	424.930	454.537
34) L7 AR-1260-4	8.181	6.993	3059059	1034657	438.304	449.630
35) L7 AR-1260-5	8.513	7.237	6457665	2415912	469.187	456.383



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021522\
Data File : P0084655.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Feb 2022 16:12
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 01:08:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 14 17:50:29 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.509 min
Delta R.T.: 0.003 min
Response: 9686508
Conc: 49.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.597 min
Delta R.T.: 0.000 min
Response: 2305333
Conc: 51.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.428 min
Delta R.T.: -0.003 min
Response: 5965657
Conc: 48.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.623 min
Delta R.T.: -0.003 min
Response: 1701243
Conc: 46.67 ng/ml



#3 AR-1016-1


R.T.: 5.698 min
Delta R.T.: 0.002 min
Response: 3002524
Conc: 463.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 835419
Conc: 486.63 ng/ml




#4 AR-1016-2

R.T.: 5.721 min
Delta R.T.: 0.002 min
Response: 4313435
Conc: 463.98 ng/ml



#4 AR-1016-2


R.T.: 4.706 min
Delta R.T.: 0.000 min
Response: 1202204
Conc: 496.75 ng/ml



#5 AR-1016-3


R.T.: 5.785 min
Delta R.T.: 0.003 min
Response: 2572652
Conc: 457.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.882 min
Delta R.T.: 0.000 min
Response: 636962
Conc: 484.70 ng/ml




#6 AR-1016-4

R.T.: 5.885 min
Delta R.T.: 0.003 min
Response: 2060679
Conc: 444.18 ng/ml



#6 AR-1016-4


R.T.: 4.926 min
Delta R.T.: 0.000 min
Response: 532079
Conc: 480.23 ng/ml



#7 AR-1016-5


R.T.: 6.184 min
Delta R.T.: 0.002 min
Response: 1960371
Conc: 437.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.139 min
Delta R.T.: 0.000 min
Response: 638815
Conc: 461.56 ng/ml




#31 AR-1260-1

R.T.: 7.326 min
Delta R.T.: 0.000 min
Response: 3128164
Conc: 452.00 ng/ml



#31 AR-1260-1


R.T.: 6.176 min
Delta R.T.: -0.002 min
Response: 1126506
Conc: 464.92 ng/ml



#32 AR-1260-2


R.T.: 7.586 min
Delta R.T.: 0.000 min
Response: 3741965
Conc: 457.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.366 min
Delta R.T.: -0.001 min
Response: 1318523
Conc: 455.20 ng/ml




#33 AR-1260-3

R.T.: 7.950 min
Delta R.T.: 0.000 min
Response: 2607422
Conc: 424.93 ng/ml



#33 AR-1260-3


R.T.: 6.519 min
Delta R.T.: -0.001 min
Response: 1236903
Conc: 454.54 ng/ml



#34 AR-1260-4


R.T.: 8.181 min
Delta R.T.: 0.000 min
Response: 3059059
Conc: 438.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.993 min
Delta R.T.: -0.002 min
Response: 1034657
Conc: 449.63 ng/ml



#35 AR-1260-5

R.T.: 8.513 min
Delta R.T.: -0.002 min
Response: 6457665
Conc: 469.19 ng/ml



#35 AR-1260-5

R.T.: 7.237 min
Delta R.T.: -0.002 min
Response: 2415912
Conc: 456.38 ng/ml