

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021524\
 Data File : P0101819.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2024 17:02
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 03:57:40 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 12 05:22:08 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.286	3.398	149.4E6	117.1E6	48.444	44.060
2) SA Decachlor...	9.843	8.278	104.4E6	82643684	49.811	43.500
Target Compounds						
26) L6 AR-1254-1	6.275	5.238	53030174	57000744	477.409	490.421
27) L6 AR-1254-2	6.491	5.383	78546328	50264205	477.899	475.451
28) L6 AR-1254-3	6.853	5.780	77687304	76578416	482.556	473.634
29) L6 AR-1254-4	7.135	6.006	53560769	44833944	540.515	533.235
30) L6 AR-1254-5	7.550	6.418	58621415	64520681	489.411	481.543



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO021524\
Data File : PO101819.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Feb 2024 17:02
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 03:57:40 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO021024.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 12 05:22:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.286 min
Delta R.T.: -0.005 min
Response: 149378551
Conc: 48.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.398 min
Delta R.T.: -0.002 min
Response: 117116826
Conc: 44.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.843 min
Delta R.T.: -0.004 min
Response: 104449716
Conc: 49.81 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.278 min
Delta R.T.: -0.006 min
Response: 82643684
Conc: 43.50 ng/ml



#26 AR-1254-1


R.T.: 6.275 min
Delta R.T.: -0.004 min
Response: 53030174
Conc: 477.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.238 min
Delta R.T.: -0.004 min
Response: 57000744
Conc: 490.42 ng/ml




#27 AR-1254-2

R.T.: 6.491 min
Delta R.T.: -0.004 min
Response: 78546328
Conc: 477.90 ng/ml



#27 AR-1254-2


R.T.: 5.383 min
Delta R.T.: -0.005 min
Response: 50264205
Conc: 475.45 ng/ml



#28 AR-1254-3


R.T.: 6.853 min
Delta R.T.: -0.003 min
Response: 77687304
Conc: 482.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.780 min
Delta R.T.: -0.004 min
Response: 76578416
Conc: 473.63 ng/ml




#29 AR-1254-4

R.T.: 7.135 min
Delta R.T.: -0.003 min
Response: 53560769
Conc: 540.52 ng/ml



#29 AR-1254-4


R.T.: 6.006 min
Delta R.T.: -0.005 min
Response: 44833944
Conc: 533.23 ng/ml



#30 AR-1254-5

R.T.: 7.550 min
Delta R.T.: -0.003 min
Response: 58621415
Conc: 489.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.418 min
Delta R.T.: -0.004 min
Response: 64520681
Conc: 481.54 ng/ml