

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021622\
 Data File : P0084710.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Feb 2022 18:15
 Operator : YP\AJ
 Sample : N1553-14 10X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LAW-2022-07B

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 17 04:49:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 14 17:50:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.596	239013	68994	1.216	1.530 #
2) SA Decachlor...	10.429	8.623	202021	46568	1.628	1.277m
Target Compounds						
26) L6 AR-1254-1	6.565	5.491	1614821	730942	249.640	285.572
27) L6 AR-1254-2	6.787	5.640	2401709	615396	243.098	274.946
28) L6 AR-1254-3	7.158	6.044	1859461	778406	179.495	213.849
29) L6 AR-1254-4	7.447	6.274	1283333	416553	175.287	185.390
30) L6 AR-1254-5	7.872	6.693	1283147	554349	159.406	169.583

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO021622\
Data File : PO084710.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Feb 2022 18:15
Operator : YP\AJ
Sample : N1553-14 10X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LAW-2022-07B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 17 04:49:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO021422.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 14 17:50:29 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

