

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021623\
 Data File : P0092503.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Feb 2023 19:38
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 21:01:41 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 04:40:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.449	3.645	137.7E6	60876731	47.555	43.355
2)	SA Decachlor...	10.374	8.683	107.2E6	50899181	51.922	45.087
Target Compounds							
21)	L5 AR-1248-1	5.635	4.732	23548357	12560595	405.695	431.155
22)	L5 AR-1248-2	5.915	4.969	37969682	19184542	415.498	425.215
23)	L5 AR-1248-3	6.122	5.011	41260815	20371201	420.712	436.779
24)	L5 AR-1248-4	6.534	5.183	41657118	23265127	425.468	433.001
25)	L5 AR-1248-5	6.573	5.577	39985973	20906244	424.121	441.600



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO021623\
Data File : PO092503.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Feb 2023 19:38
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 21:01:41 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 04:40:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: 0.000 min
Response: 137666131
Conc: 47.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.645 min
Delta R.T.: 0.000 min
Response: 60876731
Conc: 43.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.374 min
Delta R.T.: 0.002 min
Response: 107163143
Conc: 51.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: -0.010 min
Response: 50899181
Conc: 45.09 ng/ml



#21 AR-1248-1


R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 23548357
Conc: 405.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.732 min
Delta R.T.: -0.003 min
Response: 12560595
Conc: 431.15 ng/ml




#22 AR-1248-2

R.T.: 5.915 min
Delta R.T.: 0.000 min
Response: 37969682
Conc: 415.50 ng/ml



#22 AR-1248-2


R.T.: 4.969 min
Delta R.T.: -0.003 min
Response: 19184542
Conc: 425.22 ng/ml



#23 AR-1248-3


R.T.: 6.122 min
Delta R.T.: 0.000 min
Response: 41260815
Conc: 420.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.011 min
Delta R.T.: -0.003 min
Response: 20371201
Conc: 436.78 ng/ml




#24 AR-1248-4

R.T.: 6.534 min
Delta R.T.: 0.000 min
Response: 41657118
Conc: 425.47 ng/ml



#24 AR-1248-4


R.T.: 5.183 min
Delta R.T.: -0.003 min
Response: 23265127
Conc: 433.00 ng/ml



#25 AR-1248-5

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 39985973
Conc: 424.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.577 min
Delta R.T.: -0.004 min
Response: 20906244
Conc: 441.60 ng/ml