

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0021722\
 Data File : P0084730.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Feb 2022 12:15
 Operator : YP\AJ
 Sample : N1553-19 10X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 18 04:58:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0021422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 14 17:50:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.507	3.596	293097	82133	1.491	1.821
2)	SA Decachlor...	10.433	8.623	612241	59425	4.934	1.630 #
Target Compounds							
3)	L1 AR-1016-1	5.628f	0.000	23050	0	3.560	N.D. #
4)	L1 AR-1016-2	5.782f	0.000	10219	0	1.099	N.D. #
5)	L1 AR-1016-3	5.782	0.000	10219	0	1.817	N.D. #
6)	L1 AR-1016-4	5.923f	0.000	1582	0	0.341	N.D. #
7)	L1 AR-1016-5	6.180	5.089f	3336	2877	0.745	2.078 #
8)	L2 AR-1221-1	4.706	0.000	10012	0	4.630	N.D. #
9)	L2 AR-1221-2	4.819	0.000	15448	0	9.717	N.D. #
10)	L2 AR-1221-3	4.888	4.043f	11095	3050	2.239	2.344
11)	L3 AR-1232-1	4.888	4.043f	11095	3050	2.670	2.770
12)	L3 AR-1232-2	5.465f	0.000	25311	0	12.241	N.D. #
13)	L3 AR-1232-3	5.782f	0.000	10219	0	2.758	N.D. #
14)	L3 AR-1232-4	5.923f	5.027f	1582	13856	0.848	27.769 #
15)	L3 AR-1232-5	5.970	5.089f	1348	2877	0.931	5.253 #
16)	L4 AR-1242-1	5.628f	0.000	23050	0	4.636	N.D. #
17)	L4 AR-1242-2	5.782f	0.000	10219	0	1.434	N.D. #
18)	L4 AR-1242-3	5.782	0.000	10219	0	2.344	N.D. #
19)	L4 AR-1242-4	5.923f	5.027f	1582	13856	0.443	13.267 #
20)	L4 AR-1242-5	6.599	5.514	1092	1571	0.309	1.250 #
21)	L5 AR-1248-1	5.628f	0.000	23050	0	6.286	N.D. #
22)	L5 AR-1248-2	5.970	0.000	1348	0	0.261	N.D. #
23)	L5 AR-1248-3	6.180	5.027f	3336	13856	0.585	9.691 #
24)	L5 AR-1248-4	6.588	5.089f	1194	2877	0.189	1.739 #
25)	L5 AR-1248-5	6.599	5.514	1092	1571	0.181	0.960 #
26)	L6 AR-1254-1	6.569	5.514	10484	1571	1.621	0.614 #
27)	L6 AR-1254-2	6.776	5.648	351	41599	0.036	18.585 #
28)	L6 AR-1254-3	7.152	6.054	48169	14675	4.650	4.032
29)	L6 AR-1254-4	7.457	6.326f	468226	3968	63.954	1.766 #
30)	L6 AR-1254-5	7.873	6.699	12730	87622	1.581	26.805 #
31)	L7 AR-1260-1	7.324	6.209	18742	314045	2.708	129.610 #
32)	L7 AR-1260-2	7.615	6.366	228839	4408	27.995	1.522 #
33)	L7 AR-1260-3	7.904f	6.511	1708	21309	0.278	7.831 #
34)	L7 AR-1260-4	8.179	6.934f	45839	32698	6.568	14.210 #
35)	L7 AR-1260-5	8.513	7.269	2787	6338	0.202	1.197 #
36)	L8 AR-1262-1	7.904f	6.699	1708	87622	0.194	53.103 #
37)	L8 AR-1262-2	8.513	6.934f	2787	32698	0.175	11.737 #
38)	L8 AR-1262-3	8.853	7.514	105524	176880	10.030	83.905 #
39)	L8 AR-1262-4	8.974f	0.000	4938	0	1.138	N.D. #
40)	L8 AR-1262-5	9.666f	0.000	1474	0	0.266	N.D. #
41)	L9 AR-1268-1	8.853	7.514	105524	176880	5.752	28.835 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021722\
 Data File : P0084730.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Feb 2022 12:15
 Operator : YP\AJ
 Sample : N1553-19 10X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 18 04:58:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 14 17:50:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.974	7.659f	4938	175536	0.298	31.943 #
43) L9	AR-1268-3	9.188	7.749f	3093	4820	0.220	1.059 #
44) L9	AR-1268-4	9.666f	0.000	1474	0	0.243	N.D. #
45) L9	AR-1268-5	10.098	8.338f	999146	77585	21.149	5.674 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO021722\
Data File : PO084730.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Feb 2022 12:15
Operator : YP\AJ
Sample : N1553-19 10X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 18 04:58:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO021422.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 14 17:50:29 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.507 min
Delta R.T.: 0.000 min
Response: 293097
Conc: 1.49 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.596 min
Delta R.T.: 0.000 min
Response: 82133
Conc: 1.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.433 min
Delta R.T.: 0.002 min
Response: 612241
Conc: 4.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.623 min
Delta R.T.: -0.003 min
Response: 59425
Conc: 1.63 ng/ml



#3 AR-1016-1


R.T.: 5.628 min
Delta R.T.: -0.068 min
Response: 23050
Conc: 3.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.688 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.782 min
Delta R.T.: 0.064 min
Response: 10219
Conc: 1.10 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.707 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.782 min
Delta R.T.: 0.000 min
Response: 10219
Conc: 1.82 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.883 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.923 min
Delta R.T.: 0.041 min
Response: 1582
Conc: 0.34 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.926 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 6.180 min
Delta R.T.: -0.002 min
Response: 3336
Conc: 0.74 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.089 min
Delta R.T.: -0.050 min
Response: 2877
Conc: 2.08 ng/ml




#8 AR-1221-1

R.T.: 4.706 min
Delta R.T.: -0.012 min
Response: 10012
Conc: 4.63 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.813 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.819 min
Delta R.T.: 0.013 min
Response: 15448
Conc: 9.72 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 3.898 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 4.888 min
Delta R.T.: 0.004 min
Response: 11095
Conc: 2.24 ng/ml



#10 AR-1221-3


R.T.: 4.043 min
Delta R.T.: 0.068 min
Response: 3050
Conc: 2.34 ng/ml



#11 AR-1232-1


R.T.: 4.888 min
Delta R.T.: 0.004 min
Response: 11095
Conc: 2.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.043 min
Delta R.T.: 0.069 min
Response: 3050
Conc: 2.77 ng/ml




#12 AR-1232-2

R.T.: 5.465 min
Delta R.T.: 0.041 min
Response: 25311
Conc: 12.24 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T.: 4.707 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.782 min
Delta R.T.: 0.064 min
Response: 10219
Conc: 2.76 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.923 min
Delta R.T.: 0.041 min
Response: 1582
Conc: 0.85 ng/ml



#14 AR-1232-4


R.T.: 5.027 min
Delta R.T.: 0.060 min
Response: 13856
Conc: 27.77 ng/ml



#15 AR-1232-5


R.T.: 5.970 min
Delta R.T.: -0.005 min
Response: 1348
Conc: 0.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.089 min
Delta R.T.: -0.051 min
Response: 2877
Conc: 5.25 ng/ml




#16 AR-1242-1

R.T.: 5.628 min
Delta R.T.: -0.068 min
Response: 23050
Conc: 4.64 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.688 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.782 min
Delta R.T.: 0.063 min
Response: 10219
Conc: 1.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.707 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.782 min
Delta R.T.: 0.000 min
Response: 10219
Conc: 2.34 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.923 min
Delta R.T.: 0.039 min
Response: 1582
Conc: 0.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.027 min
Delta R.T.: 0.059 min
Response: 13856
Conc: 13.27 ng/ml




#20 AR-1242-5

R.T.: 6.599 min
Delta R.T.: -0.033 min
Response: 1092
Conc: 0.31 ng/ml



#20 AR-1242-5


R.T.: 5.514 min
Delta R.T.: 0.022 min
Response: 1571
Conc: 1.25 ng/ml



#21 AR-1248-1


R.T.: 5.628 min
Delta R.T.: -0.068 min
Response: 23050
Conc: 6.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.688 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.970 min
Delta R.T.: -0.005 min
Response: 1348
Conc: 0.26 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 4.926 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.180 min
Delta R.T.: -0.003 min
Response: 3336
Conc: 0.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.027 min
Delta R.T.: 0.060 min
Response: 13856
Conc: 9.69 ng/ml




#24 AR-1248-4

R.T.: 6.588 min
Delta R.T.: -0.004 min
Response: 1194
Conc: 0.19 ng/ml



#24 AR-1248-4


R.T.: 5.089 min
Delta R.T.: -0.051 min
Response: 2877
Conc: 1.74 ng/ml



#25 AR-1248-5


R.T.: 6.599 min
Delta R.T.: -0.033 min
Response: 1092
Conc: 0.18 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.514 min
Delta R.T.: -0.019 min
Response: 1571
Conc: 0.96 ng/ml




#26 AR-1254-1

R.T.: 6.569 min
Delta R.T.: 0.003 min
Response: 10484
Conc: 1.62 ng/ml



#26 AR-1254-1


R.T.: 5.514 min
Delta R.T.: 0.021 min
Response: 1571
Conc: 0.61 ng/ml



#27 AR-1254-2


R.T.: 6.776 min
Delta R.T.: -0.012 min
Response: 351
Conc: 0.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.648 min
Delta R.T.: 0.005 min
Response: 41599
Conc: 18.59 ng/ml




#28 AR-1254-3

R.T.: 7.152 min
Delta R.T.: -0.008 min
Response: 48169
Conc: 4.65 ng/ml



#28 AR-1254-3


R.T.: 6.054 min
Delta R.T.: 0.008 min
Response: 14675
Conc: 4.03 ng/ml



#29 AR-1254-4


R.T.: 7.457 min
Delta R.T.: 0.008 min
Response: 468226
Conc: 63.95 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.326 min
Delta R.T.: 0.049 min
Response: 3968
Conc: 1.77 ng/ml




#30 AR-1254-5

R.T.: 7.873 min
Delta R.T.: 0.000 min
Response: 12730
Conc: 1.58 ng/ml



#30 AR-1254-5


R.T.: 6.699 min
Delta R.T.: 0.002 min
Response: 87622
Conc: 26.80 ng/ml



#31 AR-1260-1


R.T.: 7.324 min
Delta R.T.: -0.001 min
Response: 18742
Conc: 2.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.209 min
Delta R.T.: 0.031 min
Response: 314045
Conc: 129.61 ng/ml




#32 AR-1260-2

R.T.: 7.615 min
Delta R.T.: 0.029 min
Response: 228839
Conc: 28.00 ng/ml



#32 AR-1260-2


R.T.: 6.366 min
Delta R.T.: -0.001 min
Response: 4408
Conc: 1.52 ng/ml



#33 AR-1260-3


R.T.: 7.904 min
Delta R.T.: -0.047 min
Response: 1708
Conc: 0.28 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.511 min
Delta R.T.: -0.009 min
Response: 21309
Conc: 7.83 ng/ml




#34 AR-1260-4

R.T.: 8.179 min
Delta R.T.: -0.002 min
Response: 45839
Conc: 6.57 ng/ml



#34 AR-1260-4


R.T.: 6.934 min
Delta R.T.: -0.061 min
Response: 32698
Conc: 14.21 ng/ml



#35 AR-1260-5


R.T.: 8.513 min
Delta R.T.: -0.002 min
Response: 2787
Conc: 0.20 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.269 min
Delta R.T.: 0.030 min
Response: 6338
Conc: 1.20 ng/ml




#36 AR-1262-1

R.T.: 7.904 min
Delta R.T.: -0.048 min
Response: 1708
Conc: 0.19 ng/ml



#36 AR-1262-1


R.T.: 6.699 min
Delta R.T.: 0.003 min
Response: 87622
Conc: 53.10 ng/ml



#37 AR-1262-2


R.T.: 8.513 min
Delta R.T.: -0.002 min
Response: 2787
Conc: 0.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 6.934 min
Delta R.T.: -0.061 min
Response: 32698
Conc: 11.74 ng/ml




#38 AR-1262-3

R.T.: 8.853 min
Delta R.T.: 0.009 min
Response: 105524
Conc: 10.03 ng/ml



#38 AR-1262-3


R.T.: 7.514 min
Delta R.T.: -0.011 min
Response: 176880
Conc: 83.90 ng/ml



#39 AR-1262-4


R.T.: 8.974 min
Delta R.T.: 0.036 min
Response: 4938
Conc: 1.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.588 min
Response: 0
Conc: N.D.




#40 AR-1262-5

R.T.: 9.666 min
Delta R.T.: 0.037 min
Response: 1474
Conc: 0.27 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.086 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.853 min
Delta R.T.: 0.014 min
Response: 105524
Conc: 5.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.514 min
Delta R.T.: -0.011 min
Response: 176880
Conc: 28.84 ng/ml




#42 AR-1268-2

R.T.: 8.974 min
Delta R.T.: 0.034 min
Response: 4938
Conc: 0.30 ng/ml



#42 AR-1268-2


R.T.: 7.659 min
Delta R.T.: 0.069 min
Response: 175536
Conc: 31.94 ng/ml



#43 AR-1268-3


R.T.: 9.188 min
Delta R.T.: 0.007 min
Response: 3093
Conc: 0.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.749 min
Delta R.T.: -0.048 min
Response: 4820
Conc: 1.06 ng/ml




#44 AR-1268-4

R.T.: 9.666 min
Delta R.T.: 0.036 min
Response: 1474
Conc: 0.24 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 8.086 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 10.098 min
Delta R.T.: 0.028 min
Response: 999146
Conc: 21.15 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.338 min
Delta R.T.: -0.036 min
Response: 77585
Conc: 5.67 ng/ml