

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021723\
 Data File : P0092629.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Feb 2023 14:15
 Operator : YP/AJ
 Sample : 01595-02
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 03:30:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 04:40:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.446	3.640	63722738	27063208	22.012	19.274
2)	SA Decachlor...	10.380	8.676	46916213	21917368	22.731	19.415
Target Compounds							
8)	L2 AR-1221-1	4.655	3.939f	978496	638706	25.883	33.739 #
9)	L2 AR-1221-2	4.756	3.939	876200	638706	31.661	45.447 #
10)	L2 AR-1221-3	4.756f	0.000	876200	0	10.384	N.D. #
11)	L3 AR-1232-1	4.756f	0.000	876200	0	12.663	N.D. #
12)	L3 AR-1232-2	5.331	0.000	935696	0	26.781	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021723\
Data File : P0092629.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Feb 2023 14:15
Operator : YP/AJ
Sample : 01595-02
Misc :
ALS Vial : 66 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 03:30:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 04:40:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.446 min
Delta R.T.: -0.003 min
Response: 63722738
Conc: 22.01 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.640 min
Delta R.T.: -0.006 min
Response: 27063208
Conc: 19.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.380 min
Delta R.T.: 0.009 min
Response: 46916213
Conc: 22.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.676 min
Delta R.T.: -0.017 min
Response: 21917368
Conc: 19.41 ng/ml



#8 AR-1221-1


R.T.: 4.655 min
Delta R.T.: -0.004 min
Response: 978496
Conc: 25.88 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 3.939 min
Delta R.T.: 0.063 min
Response: 638706
Conc: 33.74 ng/ml




#9 AR-1221-2

R.T.: 4.756 min
Delta R.T.: 0.010 min
Response: 876200
Conc: 31.66 ng/ml



#9 AR-1221-2


R.T.: 3.939 min
Delta R.T.: -0.022 min
Response: 638706
Conc: 45.45 ng/ml



#10 AR-1221-3


R.T.: 4.756 min
Delta R.T.: -0.068 min
Response: 876200
Conc: 10.38 ng/ml

Instrument :
ECD_O
ClientSampleId :




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 4.038 min
Response: 0
Conc: N.D.




#11 AR-1232-1

R.T.: 4.756 min
Delta R.T.: -0.067 min
Response: 876200
Conc: 12.66 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T. : 4.022 min
Response: 0
Conc: N.D.



#12 AR-1232-2

R.T.: 5.331 min
Delta R.T.: -0.031 min
Response: 935696
Conc: 26.78 ng/ml

Instrument :
ECD_O
ClientSampleId :



#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 4.753 min
Response: 0
Conc: N.D.