

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021723\
 Data File : P0092618.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Feb 2023 11:07
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 03:26:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 04:40:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.447	3.641	149.3E6	63494392	51.575	45.219
2) SA Decachlor...	10.379	8.678	112.3E6	51147706	54.409	45.307
Target Compounds						
26) L6 AR-1254-1	6.508	5.532	46440627	36491988	419.459	465.520
27) L6 AR-1254-2	6.733	5.680	77482495	32612148	476.656	471.337
28) L6 AR-1254-3	7.106	6.086	82953611	48232136	502.964	451.887
29) L6 AR-1254-4	7.399	6.315	56330708	27228358	514.975	462.325
30) L6 AR-1254-5	7.828	6.735	76000120	40364537	540.161	443.490



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO021723\
Data File : PO092618.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Feb 2023 11:07
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 03:26:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 04:40:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.447 min
Delta R.T.: -0.002 min
Response: 149303294
Conc: 51.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.641 min
Delta R.T.: -0.005 min
Response: 63494392
Conc: 45.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.379 min
Delta R.T.: 0.007 min
Response: 112297217
Conc: 54.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.678 min
Delta R.T.: -0.015 min
Response: 51147706
Conc: 45.31 ng/ml



#26 AR-1254-1


R.T.: 6.508 min
Delta R.T.: 0.002 min
Response: 46440627
Conc: 419.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.532 min
Delta R.T.: -0.007 min
Response: 36491988
Conc: 465.52 ng/ml




#27 AR-1254-2

R.T.: 6.733 min
Delta R.T.: 0.002 min
Response: 77482495
Conc: 476.66 ng/ml



#27 AR-1254-2


R.T.: 5.680 min
Delta R.T.: -0.007 min
Response: 32612148
Conc: 471.34 ng/ml



#28 AR-1254-3


R.T.: 7.106 min
Delta R.T.: 0.003 min
Response: 82953611
Conc: 502.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.086 min
Delta R.T.: -0.007 min
Response: 48232136
Conc: 451.89 ng/ml




#29 AR-1254-4

R.T.: 7.399 min
Delta R.T.: 0.003 min
Response: 56330708
Conc: 514.97 ng/ml



#29 AR-1254-4


R.T.: 6.315 min
Delta R.T.: -0.007 min
Response: 27228358
Conc: 462.33 ng/ml



#30 AR-1254-5

R.T.: 7.828 min
Delta R.T.: 0.004 min
Response: 76000120
Conc: 540.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.735 min
Delta R.T.: -0.008 min
Response: 40364537
Conc: 443.49 ng/ml