

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022120\
 Data File : P0066708.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Feb 2020 11:19
 Operator : DD\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 21 12:39:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 21 12:38:04 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.126	3.406	2086459	2629555	50.000	50.000
2) SA Decachlor...	9.596	8.288	2029971	2668680	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.272	4.460	782439	774874	500.000	500.000
4) L1 AR-1016-2	5.293	4.476	1191746	1708013	500.000	500.000
5) L1 AR-1016-3	5.354	4.649	678504	671350	500.000	500.000
6) L1 AR-1016-4	5.451	4.691	624540	570450	500.000	500.000
7) L1 AR-1016-5	5.740	4.899	544721	725250	500.000	500.000
31) L7 AR-1260-1	6.850	5.914	1094368	1424037	500.000	500.000
32) L7 AR-1260-2	7.106	6.102	1632789	1871134	500.000	500.000
33) L7 AR-1260-3	7.460	6.251	1375254	1641917	500.000	500.000
34) L7 AR-1260-4	7.684	6.716	1253843	1377777	500.000	500.000
35) L7 AR-1260-5	7.990	6.960	2914660	3587275	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022120\
Data File : PO066708.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Feb 2020 11:19
Operator : DD\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 21 12:39:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022120.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 21 12:38:04 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.126 min
Delta R.T.: 0.000 min
Response: 2086459
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.406 min
Delta R.T.: 0.000 min
Response: 2629555
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.596 min
Delta R.T.: 0.000 min
Response: 2029971
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.288 min
Delta R.T.: 0.000 min
Response: 2668680
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.272 min
Delta R.T.: 0.000 min
Response: 782439
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.460 min
Delta R.T.: 0.000 min
Response: 774874
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.293 min
Delta R.T.: 0.000 min
Response: 1191746
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.476 min
Delta R.T.: 0.000 min
Response: 1708013
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.354 min
Delta R.T.: 0.000 min
Response: 678504
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.649 min
Delta R.T.: 0.000 min
Response: 671350
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.451 min
Delta R.T.: 0.000 min
Response: 624540
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.691 min
Delta R.T.: 0.000 min
Response: 570450
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 5.740 min
Delta R.T.: 0.000 min
Response: 544721
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 725250
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 6.850 min
Delta R.T.: 0.000 min
Response: 1094368
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 5.914 min
Delta R.T.: 0.000 min
Response: 1424037
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.106 min
Delta R.T.: 0.000 min
Response: 1632789
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.102 min
Delta R.T.: 0.000 min
Response: 1871134
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.460 min
Delta R.T.: 0.000 min
Response: 1375254
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.251 min
Delta R.T.: 0.000 min
Response: 1641917
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 7.684 min
Delta R.T.: 0.000 min
Response: 1253843
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 6.716 min
Delta R.T.: 0.000 min
Response: 1377777
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.990 min
Delta R.T.: 0.000 min
Response: 2914660
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 6.960 min
Delta R.T.: 0.000 min
Response: 3587275
Conc: 500.00 ng/ml