

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022122\
 Data File : P0084785.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Feb 2022 20:56
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 00:59:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 14 17:50:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.507	3.597	9812134	2067940	49.904	45.854
2) SA Decachlor...	10.426	8.621	5967780	1467774	48.093	40.263
Target Compounds						
26) L6 AR-1254-1	6.565	5.490	3088979	1107617	477.534	432.736
27) L6 AR-1254-2	6.787	5.639	4573945	968233	462.969	432.587
28) L6 AR-1254-3	7.158	6.043	4990321	1537717	481.719	422.451
29) L6 AR-1254-4	7.447	6.273	3457521	958509	472.254	426.590
30) L6 AR-1254-5	7.871	6.692	3645366	1317836	452.867	403.145



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022122\
Data File : P0084785.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Feb 2022 20:56
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 00:59:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 14 17:50:29 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.507 min
Delta R.T.: 0.000 min
Response: 9812134
Conc: 49.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.597 min
Delta R.T.: 0.000 min
Response: 2067940
Conc: 45.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.426 min
Delta R.T.: -0.006 min
Response: 5967780
Conc: 48.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.621 min
Delta R.T.: -0.005 min
Response: 1467774
Conc: 40.26 ng/ml



#26 AR-1254-1


R.T.: 6.565 min
Delta R.T.: -0.001 min
Response: 3088979
Conc: 477.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.490 min
Delta R.T.: -0.003 min
Response: 1107617
Conc: 432.74 ng/ml




#27 AR-1254-2

R.T.: 6.787 min
Delta R.T.: -0.001 min
Response: 4573945
Conc: 462.97 ng/ml



#27 AR-1254-2


R.T.: 5.639 min
Delta R.T.: -0.003 min
Response: 968233
Conc: 432.59 ng/ml



#28 AR-1254-3


R.T.: 7.158 min
Delta R.T.: 0.000 min
Response: 4990321
Conc: 481.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.043 min
Delta R.T.: -0.003 min
Response: 1537717
Conc: 422.45 ng/ml




#29 AR-1254-4

R.T.: 7.447 min
Delta R.T.: -0.001 min
Response: 3457521
Conc: 472.25 ng/ml



#29 AR-1254-4


R.T.: 6.273 min
Delta R.T.: -0.004 min
Response: 958509
Conc: 426.59 ng/ml



#30 AR-1254-5

R.T.: 7.871 min
Delta R.T.: -0.003 min
Response: 3645366
Conc: 452.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.692 min
Delta R.T.: -0.004 min
Response: 1317836
Conc: 403.14 ng/ml