

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022123\
 Data File : P0092676.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2023 02:58
 Operator : YP/AJ
 Sample : P0022123ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO022123

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/22/2023
 Supervised By :Ankita Jodhani 02/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 05:48:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 22 05:46:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.433	3.638	165.3E6	66952416	55.035	52.525
2)	SA Decachlor...	10.261	8.672	103.8E6	54519023	53.975	46.781
Target Compounds							
3)	L1 AR-1016-1	5.609	4.725	48496500	20778105	527.659	514.332
4)	L1 AR-1016-2	5.631	4.743	72454278	29758365	543.620	515.368
5)	L1 AR-1016-3	5.694	4.919	46733424	16326316	538.268	521.006
6)	L1 AR-1016-4	5.793	4.962	35404354	14568231	542.160	521.092
7)	L1 AR-1016-5	6.091	5.175	38480001	18452625	541.438	520.191
31)	L7 AR-1260-1	7.227	6.213	66329680	34539845	543.553	506.982
32)	L7 AR-1260-2	7.486	6.401	74580503	39911131	507.422	496.514
33)	L7 AR-1260-3	7.769	6.556	80396274	37646415	511.470m	466.379
34)	L7 AR-1260-4	8.075	7.030	63617956	31114122	543.111	511.145
35)	L7 AR-1260-5	8.402	7.273	110.8E6	64185432	534.229	502.749

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022123\
Data File : PO092676.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2023 02:58
Operator : YP/AJ
Sample : PO022123ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO022123

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/22/2023
Supervised By :Ankita Jodhani 02/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 05:48:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022123.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 22 05:46:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

