

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022124\
 Data File : P0101952.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2024 15:00
 Operator : YP/AJ
 Sample : PB159182BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB159182BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/22/2024
 Supervised By :Ankita Jodhani 02/22/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 00:11:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 12 05:22:08 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.282	3.394	66492475	51172687	21.564	19.251
2) SA Decachlor...	9.832	8.268	47723758	37476167	22.759m	19.726m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022124\
Data File : PO101952.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2024 15:00
Operator : YP/AJ
Sample : PB159182BL
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB159182BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/22/2024
Supervised By :Ankita Jodhani 02/22/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 00:11:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO021024.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 12 05:22:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.282 min
Delta R.T.: -0.009 min
Response: 66492475
Conc: 21.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB159182BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/22/2024
Supervised By :Ankita Jodhani 02/22/2024




#1 Tetrachloro-m-xylene

R.T.: 3.394 min
Delta R.T.: -0.006 min
Response: 51172687
Conc: 19.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.832 min
Delta R.T.: -0.015 min
Response: 47723758
Conc: 22.76 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.268 min
Delta R.T.: -0.016 min
Response: 37476167
Conc: 19.73 ng/ml m