

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022219\
 Data File : P0054013.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Feb 2019 21:42
 Operator : SM/SJ
 Sample : PB117297BL
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB117297BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 23:06:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 19 03:40:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.059	3.137	69940321	5936068	23.321	22.246
2) SA Decachlor...	9.402	7.718	21876650	5600680	17.422	18.690

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022219\
Data File : P0054013.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Feb 2019 21:42
Operator : SM/SJ
Sample : PB117297BL
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB117297BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 23:06:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021819.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 19 03:40:07 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.059 min
Delta R.T.: -0.008 min
Response: 69940321
Conc: 23.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB117297BL




#1 Tetrachloro-m-xylene

R.T.: 3.137 min
Delta R.T.: -0.003 min
Response: 5936068
Conc: 22.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.402 min
Delta R.T.: -0.022 min
Response: 21876650
Conc: 17.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.718 min
Delta R.T.: -0.012 min
Response: 5600680
Conc: 18.69 ng/ml