

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022222\
 Data File : P0084842.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Feb 2022 18:24
 Operator : YP\AJ
 Sample : N1627-02
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 23 00:50:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 14 17:50:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.511	3.601	3992051	800057	20.303	17.740
2)	SA Decachlor...	10.435	8.624	1374429	346689	11.076	9.510
Target Compounds							
3)	L1 AR-1016-1	5.714	4.712	2140	8643	0.330	5.035 #
4)	L1 AR-1016-2	5.720	4.712	7811	8643	0.840	3.571 #
5)	L1 AR-1016-3	5.769	4.938f	88135	2552	15.670	1.942 #
6)	L1 AR-1016-4	5.879	4.938	10787	2552	2.325	2.303
7)	L1 AR-1016-5	6.192	5.095f	23891	3926	5.332	2.836 #
8)	L2 AR-1221-1	4.722	3.782	4222776	450	1952.891	0.795 #
9)	L2 AR-1221-2	4.800	3.883	47841	299623	30.092	705.665 #
10)	L2 AR-1221-3	4.869	3.957	35567	2949	7.177	2.266 #
11)	L3 AR-1232-1	4.869	3.957	35567	2949	8.560	2.678 #
12)	L3 AR-1232-2	5.400	4.712	496743	8643	240.241	8.554 #
13)	L3 AR-1232-3	5.720	4.938f	7811	2552	2.108	4.791 #
14)	L3 AR-1232-4	5.879	4.938	10787	2552	5.777	5.115
15)	L3 AR-1232-5	5.973	5.095f	47297	3926	32.673	7.168 #
16)	L4 AR-1242-1	5.714	4.712	2140	8643	0.430	6.413 #
17)	L4 AR-1242-2	5.720	4.712	7811	8643	1.096	4.597 #
18)	L4 AR-1242-3	5.769	4.938f	88135	2552	20.218	2.489 #
19)	L4 AR-1242-4	5.886	4.938	18242	2552	5.101	2.444 #
20)	L4 AR-1242-5	6.629	5.449f	43598	1416	12.333	1.127 #
21)	L5 AR-1248-1	5.714	4.712	2140	8643	0.583	8.982 #
22)	L5 AR-1248-2	5.973	4.938	47297	2552	9.154	1.862 #
23)	L5 AR-1248-3	6.192	4.938	23891	2552	4.186	1.785 #
24)	L5 AR-1248-4	6.588	5.095f	27517	3926	4.354	2.373 #
25)	L5 AR-1248-5	6.629	0.000	43598	0	7.235	N.D. #
26)	L6 AR-1254-1	6.574	5.449f	83871	1416	12.966	0.553 #
27)	L6 AR-1254-2	6.791	5.661	2115	2964	0.214	1.324 #
28)	L6 AR-1254-3	7.162	0.000	29111	0	2.810	N.D. #
29)	L6 AR-1254-4	7.454	0.000	3953	0	0.540	N.D. #
30)	L6 AR-1254-5	7.879	6.758f	10613	9587	1.318	2.933 #
31)	L7 AR-1260-1	7.320	6.177	1107	14091	0.160	5.816 #
32)	L7 AR-1260-2	7.582	6.377	814	14394	0.100	4.969 #
33)	L7 AR-1260-3	7.955	6.532	3743	2796	0.610	1.027 #
34)	L7 AR-1260-4	8.179	0.000	2548	0	0.365	N.D. #
35)	L7 AR-1260-5	8.530	7.239	23486	262305	1.706	49.551 #
36)	L8 AR-1262-1	7.955	6.758f	3743	9587	0.425	5.810 #
37)	L8 AR-1262-2	8.530	0.000	23486	0	1.475	N.D. #
38)	L8 AR-1262-3	8.843	7.518	1586	8825	0.151	4.186 #
39)	L8 AR-1262-4	8.939	7.643f	2368	4600	0.546	1.156 #
40)	L8 AR-1262-5	9.638	0.000	1352	0	0.244	N.D. #
41)	L9 AR-1268-1	8.837	7.518	2671	8825	0.146	1.439 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022222\
 Data File : P0084842.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Feb 2022 18:24
 Operator : YP\AJ
 Sample : N1627-02
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 23 00:50:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 14 17:50:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.939	7.643f	2368	4600	0.143	0.837 #
43) L9	AR-1268-3	9.183	0.000	2452	0	0.175	N.D. #
44) L9	AR-1268-4	9.638	0.000	1352	0	0.223	N.D. #
45) L9	AR-1268-5	10.069	8.373	4656	2576	0.099	0.188 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022222\
Data File : P0084842.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Feb 2022 18:24
Operator : YP\AJ
Sample : N1627-02
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 23 00:50:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 14 17:50:29 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.511 min
Delta R.T.: 0.005 min
Response: 3992051
Conc: 20.30 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.601 min
Delta R.T.: 0.004 min
Response: 800057
Conc: 17.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.435 min
Delta R.T.: 0.004 min
Response: 1374429
Conc: 11.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.624 min
Delta R.T.: -0.002 min
Response: 346689
Conc: 9.51 ng/ml



#3 AR-1016-1


R.T.: 5.714 min
Delta R.T.: 0.018 min
Response: 2140
Conc: 0.33 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.712 min
Delta R.T.: 0.024 min
Response: 8643
Conc: 5.03 ng/ml




#4 AR-1016-2

R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 7811
Conc: 0.84 ng/ml



#4 AR-1016-2


R.T.: 4.712 min
Delta R.T.: 0.006 min
Response: 8643
Conc: 3.57 ng/ml



#5 AR-1016-3


R.T.: 5.769 min
Delta R.T.: -0.013 min
Response: 88135
Conc: 15.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.938 min
Delta R.T.: 0.056 min
Response: 2552
Conc: 1.94 ng/ml




#6 AR-1016-4

R.T.: 5.879 min
Delta R.T.: -0.003 min
Response: 10787
Conc: 2.33 ng/ml



#6 AR-1016-4


R.T.: 4.938 min
Delta R.T.: 0.012 min
Response: 2552
Conc: 2.30 ng/ml



#7 AR-1016-5


R.T.: 6.192 min
Delta R.T.: 0.010 min
Response: 23891
Conc: 5.33 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.095 min
Delta R.T.: -0.043 min
Response: 3926
Conc: 2.84 ng/ml




#8 AR-1221-1

R.T.: 4.722 min
Delta R.T.: 0.004 min
Response: 4222776
Conc: 1952.89 ng/ml



#8 AR-1221-1


R.T.: 3.782 min
Delta R.T.: -0.031 min
Response: 450
Conc: 0.79 ng/ml



#9 AR-1221-2


R.T.: 4.800 min
Delta R.T.: -0.006 min
Response: 47841
Conc: 30.09 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.883 min
Delta R.T.: -0.015 min
Response: 299623
Conc: 705.67 ng/ml




#10 AR-1221-3

R.T.: 4.869 min
Delta R.T.: -0.015 min
Response: 35567
Conc: 7.18 ng/ml



#10 AR-1221-3


R.T.: 3.957 min
Delta R.T.: -0.018 min
Response: 2949
Conc: 2.27 ng/ml



#11 AR-1232-1


R.T.: 4.869 min
Delta R.T.: -0.014 min
Response: 35567
Conc: 8.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.957 min
Delta R.T.: -0.018 min
Response: 2949
Conc: 2.68 ng/ml




#12 AR-1232-2

R.T.: 5.400 min
Delta R.T.: -0.023 min
Response: 496743
Conc: 240.24 ng/ml



#12 AR-1232-2


R.T.: 4.712 min
Delta R.T.: 0.005 min
Response: 8643
Conc: 8.55 ng/ml



#13 AR-1232-3


R.T.: 5.720 min
Delta R.T.: 0.001 min
Response: 7811
Conc: 2.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.938 min
Delta R.T.: 0.056 min
Response: 2552
Conc: 4.79 ng/ml




#14 AR-1232-4

R.T.: 5.879 min
Delta R.T.: -0.003 min
Response: 10787
Conc: 5.78 ng/ml



#14 AR-1232-4


R.T.: 4.938 min
Delta R.T.: -0.029 min
Response: 2552
Conc: 5.11 ng/ml



#15 AR-1232-5


R.T.: 5.973 min
Delta R.T.: -0.002 min
Response: 47297
Conc: 32.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.095 min
Delta R.T.: -0.044 min
Response: 3926
Conc: 7.17 ng/ml




#16 AR-1242-1

R.T.: 5.714 min
Delta R.T.: 0.018 min
Response: 2140
Conc: 0.43 ng/ml



#16 AR-1242-1


R.T.: 4.712 min
Delta R.T.: 0.024 min
Response: 8643
Conc: 6.41 ng/ml



#17 AR-1242-2


R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 7811
Conc: 1.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.712 min
Delta R.T.: 0.005 min
Response: 8643
Conc: 4.60 ng/ml




#18 AR-1242-3

R.T.: 5.769 min
Delta R.T.: -0.014 min
Response: 88135
Conc: 20.22 ng/ml



#18 AR-1242-3


R.T.: 4.938 min
Delta R.T.: 0.056 min
Response: 2552
Conc: 2.49 ng/ml



#19 AR-1242-4


R.T.: 5.886 min
Delta R.T.: 0.003 min
Response: 18242
Conc: 5.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.938 min
Delta R.T.: -0.030 min
Response: 2552
Conc: 2.44 ng/ml




#20 AR-1242-5

R.T.: 6.629 min
Delta R.T.: -0.002 min
Response: 43598
Conc: 12.33 ng/ml



#20 AR-1242-5


R.T.: 5.449 min
Delta R.T.: -0.043 min
Response: 1416
Conc: 1.13 ng/ml



#21 AR-1248-1


R.T.: 5.714 min
Delta R.T.: 0.017 min
Response: 2140
Conc: 0.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.712 min
Delta R.T.: 0.024 min
Response: 8643
Conc: 8.98 ng/ml




#22 AR-1248-2

R.T.: 5.973 min
Delta R.T.: -0.003 min
Response: 47297
Conc: 9.15 ng/ml



#22 AR-1248-2


R.T.: 4.938 min
Delta R.T.: 0.012 min
Response: 2552
Conc: 1.86 ng/ml



#23 AR-1248-3


R.T.: 6.192 min
Delta R.T.: 0.010 min
Response: 23891
Conc: 4.19 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.938 min
Delta R.T.: -0.029 min
Response: 2552
Conc: 1.78 ng/ml




#24 AR-1248-4

R.T.: 6.588 min
Delta R.T.: -0.005 min
Response: 27517
Conc: 4.35 ng/ml



#24 AR-1248-4


R.T.: 5.095 min
Delta R.T.: -0.044 min
Response: 3926
Conc: 2.37 ng/ml



#25 AR-1248-5


R.T.: 6.629 min
Delta R.T.: -0.003 min
Response: 43598
Conc: 7.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.534 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.574 min
Delta R.T.: 0.008 min
Response: 83871
Conc: 12.97 ng/ml



#26 AR-1254-1


R.T.: 5.449 min
Delta R.T.: -0.043 min
Response: 1416
Conc: 0.55 ng/ml



#27 AR-1254-2


R.T.: 6.791 min
Delta R.T.: 0.003 min
Response: 2115
Conc: 0.21 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.661 min
Delta R.T.: 0.019 min
Response: 2964
Conc: 1.32 ng/ml




#28 AR-1254-3

R.T.: 7.162 min
Delta R.T.: 0.003 min
Response: 29111
Conc: 2.81 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.046 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.454 min
Delta R.T.: 0.006 min
Response: 3953
Conc: 0.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.277 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.879 min
Delta R.T.: 0.005 min
Response: 10613
Conc: 1.32 ng/ml



#30 AR-1254-5


R.T.: 6.758 min
Delta R.T.: 0.061 min
Response: 9587
Conc: 2.93 ng/ml



#31 AR-1260-1


R.T.: 7.320 min
Delta R.T.: -0.006 min
Response: 1107
Conc: 0.16 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.177 min
Delta R.T.: 0.000 min
Response: 14091
Conc: 5.82 ng/ml




#32 AR-1260-2

R.T.: 7.582 min
Delta R.T.: -0.003 min
Response: 814
Conc: 0.10 ng/ml



#32 AR-1260-2


R.T.: 6.377 min
Delta R.T.: 0.010 min
Response: 14394
Conc: 4.97 ng/ml



#33 AR-1260-3


R.T.: 7.955 min
Delta R.T.: 0.004 min
Response: 3743
Conc: 0.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.532 min
Delta R.T.: 0.011 min
Response: 2796
Conc: 1.03 ng/ml




#34 AR-1260-4

R.T.: 8.179 min
Delta R.T.: -0.002 min
Response: 2548
Conc: 0.37 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 6.995 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.530 min
Delta R.T.: 0.016 min
Response: 23486
Conc: 1.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.239 min
Delta R.T.: 0.000 min
Response: 262305
Conc: 49.55 ng/ml




#36 AR-1262-1

R.T.: 7.955 min
Delta R.T.: 0.004 min
Response: 3743
Conc: 0.42 ng/ml



#36 AR-1262-1


R.T.: 6.758 min
Delta R.T.: 0.062 min
Response: 9587
Conc: 5.81 ng/ml



#37 AR-1262-2


R.T.: 8.530 min
Delta R.T.: 0.015 min
Response: 23486
Conc: 1.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 6.996 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.843 min
Delta R.T.: 0.000 min
Response: 1586
Conc: 0.15 ng/ml



#38 AR-1262-3


R.T.: 7.518 min
Delta R.T.: -0.006 min
Response: 8825
Conc: 4.19 ng/ml



#39 AR-1262-4


R.T.: 8.939 min
Delta R.T.: 0.001 min
Response: 2368
Conc: 0.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.643 min
Delta R.T.: 0.055 min
Response: 4600
Conc: 1.16 ng/ml




#40 AR-1262-5

R.T.: 9.638 min
Delta R.T.: 0.009 min
Response: 1352
Conc: 0.24 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.086 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.837 min
Delta R.T.: -0.003 min
Response: 2671
Conc: 0.15 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.518 min
Delta R.T.: -0.006 min
Response: 8825
Conc: 1.44 ng/ml




#42 AR-1268-2

R.T.: 8.939 min
Delta R.T.: -0.001 min
Response: 2368
Conc: 0.14 ng/ml



#42 AR-1268-2


R.T.: 7.643 min
Delta R.T.: 0.053 min
Response: 4600
Conc: 0.84 ng/ml



#43 AR-1268-3


R.T.: 9.183 min
Delta R.T.: 0.002 min
Response: 2452
Conc: 0.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.798 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.638 min
Delta R.T.: 0.008 min
Response: 1352
Conc: 0.22 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 8.086 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 10.069 min
Delta R.T.: 0.000 min
Response: 4656
Conc: 0.10 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.373 min
Delta R.T.: -0.001 min
Response: 2576
Conc: 0.19 ng/ml