

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022222\
 Data File : P0084834.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Feb 2022 15:45
 Operator : YP\AJ
 Sample : N1626-07
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CC-A4

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 23 00:44:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 14 17:50:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.598	3762944	844658	19.138	18.729
2) SA Decachlor...	10.434	8.625	2737056	593879	22.057	16.291 #
Target Compounds						
26) L6 AR-1254-1	6.568	5.492	2229074	447176	344.599	174.707 #
27) L6 AR-1254-2	6.790	5.642	2213031	474459	224.000	211.979
28) L6 AR-1254-3	7.161	6.046	2971188	865828	286.811	237.866
29) L6 AR-1254-4	7.451	6.276	3135131	703550	428.220	313.119 #
30) L6 AR-1254-5	7.876	6.696	2821555	1110341	350.524	339.669



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022222\
Data File : P0084834.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Feb 2022 15:45
Operator : YP\AJ
Sample : N1626-07
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CC-A4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 23 00:44:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021422.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 14 17:50:29 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.510 min
Delta R.T.: 0.004 min
Response: 3762944
Conc: 19.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
CC-A4




#1 Tetrachloro-m-xylene

R.T.: 3.598 min
Delta R.T.: 0.002 min
Response: 844658
Conc: 18.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.003 min
Response: 2737056
Conc: 22.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.625 min
Delta R.T.: -0.001 min
Response: 593879
Conc: 16.29 ng/ml



#26 AR-1254-1


R.T.: 6.568 min
Delta R.T.: 0.002 min
Response: 2229074
Conc: 344.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
CC-A4




#26 AR-1254-1

R.T.: 5.492 min
Delta R.T.: 0.000 min
Response: 447176
Conc: 174.71 ng/ml




#27 AR-1254-2

R.T.: 6.790 min
Delta R.T.: 0.002 min
Response: 2213031
Conc: 224.00 ng/ml



#27 AR-1254-2


R.T.: 5.642 min
Delta R.T.: 0.000 min
Response: 474459
Conc: 211.98 ng/ml



#28 AR-1254-3


R.T.: 7.161 min
Delta R.T.: 0.002 min
Response: 2971188
Conc: 286.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
CC-A4




#28 AR-1254-3

R.T.: 6.046 min
Delta R.T.: 0.000 min
Response: 865828
Conc: 237.87 ng/ml




#29 AR-1254-4

R.T.: 7.451 min
Delta R.T.: 0.002 min
Response: 3135131
Conc: 428.22 ng/ml



#29 AR-1254-4


R.T.: 6.276 min
Delta R.T.: -0.001 min
Response: 703550
Conc: 313.12 ng/ml



#30 AR-1254-5

R.T.: 7.876 min
Delta R.T.: 0.002 min
Response: 2821555
Conc: 350.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
CC-A4



#30 AR-1254-5

R.T.: 6.696 min
Delta R.T.: -0.001 min
Response: 1110341
Conc: 339.67 ng/ml