

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022321\
 Data File : P0075677.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Feb 2021 18:54
 Operator : AJ/MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 24 01:07:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 23 07:07:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.944	3.963	1510738	1033044	22.935	24.945
2)	SA Decachlor...	10.991	9.255	1337454	1134569	22.756	23.331
Target Compounds							
9)	L2 AR-1221-2	5.297	0.000	22779	0	38.226	N.D. #
24)	L5 AR-1248-4	7.211	0.000	459495	0	170.311	N.D. #
27)	L6 AR-1254-2	7.406	0.000	123751	0	30.278	N.D. #
28)	L6 AR-1254-3	7.803f	0.000	157001	0	35.438	N.D. #
29)	L6 AR-1254-4	8.084	0.000	382596	0	117.529	N.D. #
30)	L6 AR-1254-5	8.518	0.000	105720	0	30.202	N.D. #
32)	L7 AR-1260-2	8.247f	0.000	247808	0	69.405	N.D. #
33)	L7 AR-1260-3	0.000	7.155	0	30012	N.D.	10.645 #
34)	L7 AR-1260-4	8.812	0.000	219812	0	67.405	N.D. #
39)	L8 AR-1262-4	0.000	8.218f	0	33385	N.D.	6.609 #
42)	L9 AR-1268-2	0.000	8.218f	0	33385	N.D.	4.759 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022321\
Data File : PO075677.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Feb 2021 18:54
Operator : AJ/MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 24 01:07:16 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022221.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 23 07:07:50 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.944 min
Delta R.T.: 0.002 min
Response: 1510738
Conc: 22.94 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.963 min
Delta R.T.: -0.002 min
Response: 1033044
Conc: 24.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.991 min
Delta R.T.: 0.004 min
Response: 1337454
Conc: 22.76 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.255 min
Delta R.T.: -0.004 min
Response: 1134569
Conc: 23.33 ng/ml



#9 AR-1221-2


R.T.: 5.297 min
Delta R.T.: 0.005 min
Response: 22779
Conc: 38.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 4.320 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 7.211 min
Delta R.T.: 0.009 min
Response: 459495
Conc: 170.31 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.709 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 7.406 min
Delta R.T.: 0.004 min
Response: 123751
Conc: 30.28 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 6.248 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.803 min
Delta R.T.: 0.018 min
Response: 157001
Conc: 35.44 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.668 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 8.084 min
Delta R.T.: 0.002 min
Response: 382596
Conc: 117.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.907 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 8.518 min
Delta R.T.: 0.006 min
Response: 105720
Conc: 30.20 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 7.336 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 8.247 min
Delta R.T.: 0.027 min
Response: 247808
Conc: 69.41 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.004 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 8.586 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.155 min
Delta R.T.: -0.003 min
Response: 30012
Conc: 10.65 ng/ml



#34 AR-1260-4


R.T.: 8.812 min
Delta R.T.: -0.007 min
Response: 219812
Conc: 67.41 ng/ml

Instrument :
ECD_O
ClientSampleId :




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.641 min
Response: 0
Conc: N.D.




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 9.555 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.218 min
Delta R.T.: -0.019 min
Response: 33385
Conc: 6.61 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T.: 9.559 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#42 AR-1268-2

R.T.: 8.218 min
Delta R.T.: -0.020 min
Response: 33385
Conc: 4.76 ng/ml