

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022324\
 Data File : P0102040.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Feb 2024 23:44
 Operator : YP/AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 24 03:46:58 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 24 03:45:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.282	3.392	224.8E6	157.0E6	50.000	50.000
2) SA Decachlor...	9.841	8.272	126.5E6	103.7E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.272	5.232	68868969	73420027	500.000	500.000
27) L6 AR-1254-2	6.487	5.377	100.5E6	64931736	500.000	500.000
28) L6 AR-1254-3	6.851	5.774	99481525	99559760	500.000	500.000
29) L6 AR-1254-4	7.133	6.000	63768716	54808858	500.000	500.000
30) L6 AR-1254-5	7.548	6.412	72249558	83865658	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022324\
Data File : PO102040.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Feb 2024 23:44
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 24 03:46:58 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022324.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 24 03:45:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.282 min
Delta R.T.: 0.000 min
Response: 224828475
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.392 min
Delta R.T.: 0.000 min
Response: 157016552
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.841 min
Delta R.T.: 0.000 min
Response: 126533059
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.272 min
Delta R.T.: 0.000 min
Response: 103728748
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.272 min
Delta R.T.: 0.000 min
Response: 68868969
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.232 min
Delta R.T.: 0.000 min
Response: 73420027
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.487 min
Delta R.T.: 0.000 min
Response: 100538075
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.377 min
Delta R.T.: 0.000 min
Response: 64931736
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 6.851 min
Delta R.T.: 0.000 min
Response: 99481525
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 5.774 min
Delta R.T.: 0.000 min
Response: 99559760
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.133 min
Delta R.T.: 0.000 min
Response: 63768716
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.000 min
Delta R.T.: 0.000 min
Response: 54808858
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.548 min
Delta R.T.: 0.000 min
Response: 72249558
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.412 min
Delta R.T.: 0.000 min
Response: 83865658
Conc: 500.00 ng/ml