

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022421\
 Data File : P0075704.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Feb 2021 11:36
 Operator : AJ/MA
 Sample : M1466-01 5X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 JEC402U-END-1-1

Manual Integrations
 APPROVED

Ankita
 2/25/2021 8:53:34 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 24 15:33:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 23 07:07:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.944	3.964	277739	179130	4.216	4.326
2)	SA Decachlor...	11.008	9.256	196586	157561	3.345m	3.240
Target Compounds							
26)	L6 AR-1254-1	7.173	6.084	32575	29017	12.413m	10.663
27)	L6 AR-1254-2	7.404	6.245	62039	25229	15.179m	10.556 #
28)	L6 AR-1254-3	7.787	6.663	66095	53572	14.919	13.706
29)	L6 AR-1254-4	8.083	6.902	41277	24796	12.680m	9.540
30)	L6 AR-1254-5	8.515	7.331	55970	40526	15.989m	11.409 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO022421\
Data File : PO075704.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Feb 2021 11:36
Operator : AJ/MA
Sample : M1466-01 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
JEC402U-END-1-1


Manual Integrations
APPROVED

Ankita
2/25/2021 8:53:34 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 24 15:33:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022221.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 23 07:07:50 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.944 min
Delta R.T.: 0.002 min
Response: 277739
Conc: 4.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
JEC402U-END-1-1


Manual Integrations
APPROVED

Ankita
2/25/2021 8:53:34 AM




#1 Tetrachloro-m-xylene

R.T.: 3.964 min
Delta R.T.: 0.000 min
Response: 179130
Conc: 4.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.008 min
Delta R.T.: 0.013 min
Response: 196586
Conc: 3.34 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.256 min
Delta R.T.: -0.003 min
Response: 157561
Conc: 3.24 ng/ml




#26 AR-1254-1

R.T.: 7.173 min
Delta R.T.: 0.001 min
Response: 32575
Conc: 12.41 ng/ml m

Instrument :
ECD_O
ClientSampleId :
JEC402U-END-1-1


Manual Integrations
APPROVED

Ankita
2/25/2021 8:53:34 AM




#26 AR-1254-1

R.T.: 6.084 min
Delta R.T.: -0.004 min
Response: 29017
Conc: 10.66 ng/ml




#27 AR-1254-2

R.T.: 7.404 min
Delta R.T.: 0.002 min
Response: 62039
Conc: 15.18 ng/ml m



#27 AR-1254-2

R.T.: 6.245 min
Delta R.T.: -0.003 min
Response: 25229
Conc: 10.56 ng/ml




#28 AR-1254-3

R.T.: 7.787 min
Delta R.T.: 0.002 min
Response: 66095
Conc: 14.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
JEC402U-END-1-1


Manual Integrations
APPROVED

Ankita
2/25/2021 8:53:34 AM




#28 AR-1254-3

R.T.: 6.663 min
Delta R.T.: -0.005 min
Response: 53572
Conc: 13.71 ng/ml




#29 AR-1254-4

R.T.: 8.083 min
Delta R.T.: 0.001 min
Response: 41277
Conc: 12.68 ng/ml m



#29 AR-1254-4

R.T.: 6.902 min
Delta R.T.: -0.005 min
Response: 24796
Conc: 9.54 ng/ml




#30 AR-1254-5

R.T.: 8.515 min
Delta R.T.: 0.003 min
Response: 55970
Conc: 15.99 ng/ml m

Instrument :
ECD_O
ClientSampleId :
JEC402U-END-1-1

Manual Integrations
APPROVED

Ankita
2/25/2021 8:53:34 AM



#30 AR-1254-5

R.T.: 7.331 min
Delta R.T.: -0.005 min
Response: 40526
Conc: 11.41 ng/ml